



HAL
open science

In-mold patterning and actionable axo-somatic compartmentalization for on-chip neuron culture

Ayako Yamada, Maéva Vignes, Cécile Bureau, Alexandre Mamane, Bastien Venzac, Stéphanie Descroix, Jean-Louis Viovy, Catherine Villard, Jean-Michel Peyrin, Laurent Malaquin

► **To cite this version:**

Ayako Yamada, Maéva Vignes, Cécile Bureau, Alexandre Mamane, Bastien Venzac, et al.. In-mold patterning and actionable axo-somatic compartmentalization for on-chip neuron culture. *Lab on a Chip*, 2016, 16 (11), pp.2059-2068. 10.1039/C6LC00414H . hal-01324932

HAL Id: hal-01324932

<https://hal.sorbonne-universite.fr/hal-01324932v1>

Submitted on 1 Jun 2016

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

In-mold patterning and actionable axo-somatic compartmentalization for on-chip neuron culture

Ayako Yamada,^{‡ab} Maéva Vignes,^{‡abc} Cécile Bureau,^{ab} Alexandre Mamane,^{ab} Bastien Venzac,^{ab} Stéphanie

^a Institut Curie, PSL Research University, CNRS, UMR 168, F-75005, Paris, France

^b Sorbonne Universités, UPMC Univ Paris 06, CNRS, UMR 168, F-75005, Paris, France

^c UPMC Univ Paris 06, CNRS, UMR 8256, B2A, Sorbonne Universités, Biological Adaptation and Ageing, Institut de Biologie Paris Seine, Paris, F-75005, France

^d CNRS, Inst NEEL and CRETA, Univ. Grenoble Alpes, F-38042 Grenoble, France

^e LAAS-CNRS, Université de Toulouse, CNRS, Toulouse, France* Corresponding authors.

§ Equal contribution.

Electronic Supplementary Information (ESI) available: Fig. S1 – S4 and text S1. See DOI: 10.1039/x0xx00000x

Descroix,^{ab} Jean-Louis Viovy,^{ab} Catherine Villard,^{*abd} Jean-Michel Peyrin^{*c} and Laurent Malaquin^{*abe}

Oriented neuronal networks with controlled connectivity are required for many applications ranging from studies of neurodegeneration to neuronal computation. To build such networks in vitro, an efficient, directed and long lasting guidance of axons toward their target is a pre-requisite. The best guidance achieved so far, however, relied on confining axons in enclosed microchannels, making them poorly accessible for further investigation. Here we describe a method providing accessible and highly regular arrays of axons, emanating from soma positioned in distinct compartments. This method combines the use of a novel removable partition, allowing soma positioning outside of the axon guidance patterns, and in-mold patterning (iMP), a hybrid method combining chemical and mechanical cell positioning clues applied here for the first time to neurons. The axon guidance efficiency of iMP is compared to that of conventional patterning methods, e.g. micro-contact printing (chemical constraints by a poly-L-lysine motif) and micro-grooves (physical constraints by homogeneously coated microstructures), using guiding tracks of different widths and spacing. We show that iMP provides a gain of 10 to 100 in axon confinement efficiency on the tracks, yielding mm-long, highly regular, and fully accessible on-chip axon arrays. iMP also allows well-defined axon guidance from small populations of several neurons confined at predefined positions in μm -sized wells. iMP will thus open new routes for the construction of complex and accurately controlled neuronal networks.

Introduction

The complex ramified morphology of cultured neurons raises difficulties in analyzing their behaviors or in constructing artificial neuronal networks of controlled axo-dendritic polarity. In the past decades, strong efforts have thus been made to control neuronal morphologies.¹ Methods derived from soft lithography² are particularly suited to build substrates with micro-scale features specifically designed to direct the shape of neurons. Micro-contact printing (μCP) techniques, in which micro-patterns of biomolecules are printed on flat substrates, were applied for controlling the morphology of neurons.^{3,4} They provide chemical constraints on neuronal morphologies, by making a contrast of cell adhesion properties inside and outside the micro-patterns. Other groups used physical constraints on neurons, such as topographical barriers in the form of micro-grooves⁵⁻⁷ or micro-pillars.^{8,9} However, by applying a sole chemical or physical constraint, achieving a long term and efficient confinement of neurons on μm -sized motifs remains challenging.^{6,7,10,11} To further improve neuron confinement, the combination of chemical and physical constraints, referred to as hybrid constraint, has also been explored. For instance, Zhang et al. used topographical barriers with a partial chemical treatment only at the bottom of the structures.¹² However, the authors observed that axons climbed over the barriers made of a photoresist, to which axons could adhere even without surface treatment. In a different context, Biancardo et al. had proposed another method to prepare substrates with hybrid constraints, referred to as in-mold patterning (iMP).^{13,14} It consists in transferring both protein micro-patterns and microstructures simultaneously from a mold to the final substrate, thanks to a step in which the protein is deposited on the protruding part of the microstructures of the mold. Here we propose to adapt this method by replacing polymer melts with polydimethylsiloxane (PDMS) for easy handling and versatility, and apply it to neurons for improved control over their morphology.

Besides the difficulties discussed above, regarding the quality of axon guidance, the preparation of well-controlled neuron arrays faces another challenge, associated with the relative positioning of soma with regards to the axon guidance tracks. Axon guidance in combination with soma positioning has been achieved by several strategies. For instance, micro-patterns of cell adhesive molecules,^{10,15-18} cell repellent layers,¹⁹ or a modifiable microstructure in agarose gel²⁰ were proposed, but they are limited to single or a few neurons. For larger neuron populations, Park et al. combined PDMS compartments with a micro-grooved substrate to obtain an open array of axons with soma positioning, but long-term axon confinement was not demonstrated.⁷ In addition to these strategies, stencil seeding techniques,²¹ or laminar structures of hydrogel in a microfluidic channel²² provide soma positioning but no axon guidance.




Fig. 1. Scheme of three different substrates for axon guidance. Micro-contact printing (μ CP) substrate consists in a micro-pattern of poly-L-lysine-FITC (PLL-FITC) on a flat surface, which works as a chemical constraint for axons (left). Micro-groove (μ Gro) substrate consists in 5 μ m-high micro-structures with homogeneous PLL-FITC coating, providing a topographical constraint for axons (middle). In-mold patterning (iMP) substrate is a hybrid substrate with 5 μ m-high micro-structures with PLL-FITC coating at the bottom of the structure only (right). The surface areas coated with PLL-FITC are indicated with green color.

A more satisfactory solution to the problem of somato-axonal segregation was provided by microfluidics.²³ In the last years the development of various microfluidic platforms has considerably increased the ability of creating deterministic neuron arrays.²⁴⁻²⁷ In such platforms, axons were separated from somas and guided by narrow and closed microchannels with their height smaller than that of somas, thus allowing a full exclusion of soma from the guiding tracks. By using funnel-shaped microchannels, this technique also permitted the mastering of a directional connectivity between two populations.²⁶ However, all these approaches share the limitations of strongly confining axons in narrow microchannels, resulting in poor accessibility of the guided axons, and nutrition problems for long term viability of the axons.

Combining a full spatial separation between somato-dendritic and axonal compartments with a long term confinement of axons in high-density neuron cultures, while keeping all cellular compartments in an open microenvironment to allow physical, biochemical, and cellular addressing, thus remains an unmet challenge. Here, we address this challenge using a novel microfilament-integrated microfluidic chip. The microfilament serves as a removable wall, thus allowing post-seeding reconfiguration of the partition of the microfluidic platform.²⁸ The partition separates the seeding chamber from axon guiding tracks during soma positioning, enabling a well-defined control of soma positioning without the high confinement limit imposed by microchannels in earlier methods.²⁴⁻²⁷ We combine this approach with a modified and optimized iMP, and compare it regarding axon guidance efficiency with two more conventional techniques: μ CP and topographical constraint in the form of micro-grooves (μ Gro). All the substrates are made in a thin layer of PDMS with (μ Gro, iMP) or without (μ CP) 5 μ m-high microstructures. Substrates are coated with fluorescein isothiocyanate-labeled poly-L-lysine (PLL-FITC) partially (μ CP, iMP) or homogeneously (μ Gro) for cell adhesion, as depicted in Fig. 1 (green color). Overall, we show that iMP technique achieves a much better long-term confinement of axons originating from a large or small neuronal population, thus paving the way for reconstituting highly deterministic neuronal networks. Systematic studies, varying geometrical parameters, allow us to explore and discuss the mechanisms of confinement, and provide guidelines for optimizing micro-pattern geometries.

Materials and Methods

All reagents were purchased from Sigma-Aldrich unless otherwise specified.

Preparation of axon guiding substrates

Micro-contact printing (μ CP). μ CP consists in a micro-patterning of PLL-FITC on a flat substrate, performed by transferring onto the substrate PLL-FITC from a PDMS stamp with 5 μ m-high microstructures (Fig. 2, top). In contrast with the substrates previously used for micro-contact printing to guide neurites, such as glass, polystyrene, or agarose-coated glass substrates,^{4, 11, 29, 30} a PDMS-coated glass slide was chosen here for μ CP. This allows a more direct comparison with the other patterning methods, i.e. μ Gro and iMP, without any bias due to the physico-chemical property of the substrate material. The thickness of the PDMS layer was kept small as compared to iMP and μ Gro, in order to keep a substrate flat and solid enough to allow a good contact with a PDMS stamp, and to obtain a uniform and high resolution micro-pattern. The substrate was prepared by spreading non-reticulated PDMS (Dow Corning), mixed with its linker at the ratio of 10:1 (w/w) and degassed prior to use, on a glass cover slip (0.13-0.16 mm thick) by spin-coater (SPIN150, SPS-Europe) at 3000 rpm for 60 s and incubated at 66 °C overnight. The thickness of the PDMS layer was measured to be 28 ± 1 μ m by profilometer (Wyko NT1100, Veeco). To better transfer PLL-FITC from a PDMS stamp to the substrate, which is hydrophobic by nature, the substrate surface was turned to a high energy surface by plasma treatment using an air plasma cleaner (Harrick) immediately before printing.




Fig. 2. Scheme of the preparation of three different substrates and microfluidic chip assembly. The thickness of all the SU-8 layers is 5 μm .

The mold for the PDMS stamp was fabricated by photolithography as follows. A 5 μm -thick layer of SU-8 2005 (MicroChem) was formed on a 2 inch silicon wafer (Neyco) by spin-coating at 3000 rpm for 30 s. After a soft bake, the wafer was exposed to a UV light (MJB4 Mask Aligner, SUSS MicroTec) through a hard chromium mask or a flexible transparency mask (Selba S. A.) depending on the required resolutions prior to a post exposure bake, and developed using SU-8 developer (MicroChem). The obtained mold was exposed to a vapor of trimethylchlorosilane (TMCS, ABCR GmbH & Co. KG) for 5 min once, and used repeatedly without further treatment. PDMS was mixed with its linker, poured on the mold, degassed under vacuum, and reticulated at 66 $^\circ\text{C}$ for at least 4 h. A PDMS block, about 3 mm thick, was cut off from the mold and used as a stamp. The stamp was soaked in 96 % ethanol for 24 h, dried at 66 $^\circ\text{C}$ overnight, and plasma treated prior to the exposure to the vapor of (tridecafluoro-1,1,2,2-tetrahydrooctyl)trichlorosilane (ABCR GmbH & Co. KG) in a depressurized desiccator for 20 min. This silanization step prevented the stamp surface from binding to the plasma treated substrate during the following stamping process. The stamp was incubated in 10 % (w/v) sodium dodecyl sulfate (SDS, Euromedex) in MilliQ water for 20 min to decorate the hydrophobic stamp surface with negative charges and to enhance PLL-FITC transfer from the stamp to the substrate,³¹ rinsed 3 times with MilliQ water, and incubated with 0.01 mg/mL PLL-FITC in Dulbecco's phosphate buffered saline (DPBS) without CaCl_2 and MgCl_2 for 1 h at room temperature. After being rinsed with water and dried by pressurized air flow, the stamp was carefully laid down on the plasma treated substrate with its microstructured side down. A 30 g flat weight was put on the stamp for 3 min and removed. The substrate, together with the stamp on it, was aligned under the microscope on a printed paper reproducing the exact shape of a microfluidic channel (see below). The substrate was maintained with Scotch removable tapes (3M) on this paper template. This step was essential in order to have a reference in terms of positioning for the final mounting of the channel on the micro-printed substrate. The stamp was then removed from the substrate and discarded. The substrate was dried at room temperature overnight. The retention of the PLL-FITC micro-pattern was strongly improved by the drying process, while the non-covered substrate surface returned back to a hydrophobic state during drying. Indeed, we could not obtain a clear PLL-FITC micro-pattern when the substrate was immersed in a buffer immediately after printing. This implies that the drying process promotes the physisorption of PLL-FITC on the substrate.

In-mold patterning (iMP). In the iMP approach, we prepared 5 μm -high microstructures with PLL-FITC coating at the bottom of the structures only. To achieve such features, a thin PDMS layer was deposited by spin-coating on a stamp inked with PLL-FITC, and then reticulated (Fig. 2, middle). More precisely, the PDMS stamp, about 5 mm thick and with 1 cm of margin around the microstructures, was first prepared using the same procedure as above for μCP , including the silanization process, which was necessary to avoid the fusion of the PDMS stamp with the patterned layer. An inker was prepared by reticulating degassed PDMS mixture in a plastic Petri dish at 66 $^{\circ}\text{C}$ for at least 4 h. Both the stamp and the inker were incubated with 96 % ethanol for 24 h and dried at 66 $^{\circ}\text{C}$ overnight, incubated with 10 % SDS in water for 20 min, and rinsed 3 times with water prior to use. The inker was incubated with 0.01 mg/mL PLL-FITC in DPBS for 1 h at room temperature, rinsed with water, and dried with a pressurized air




Fig. 3. Scheme of the microfluidic chip for cell culture and axon guidance. An 80 μm -high microfluidic channel with 4 reservoirs, equipped with a removable partition made of a fishing line, was aligned and bound on an axon guiding substrate (left). Cells were settled on a cell seeding area next to axon guiding tracks, separated by the partition. The latter was removed 1 hour after cell seeding (right).

flow immediately before use. The stamp was dried with a pressurized air flow, put on the inker with its microstructured surface down, and pressed lightly to make a good contact with the inker. The stamp and inker were kept in contact for at least 3 min before removal of the stamp. Non-reticulated, degassed PDMS mixture was spread on the inked side of the stamp by spin-coating at 800 rpm for 30 s, prior to incubation at 66 $^{\circ}\text{C}$ overnight. The edge of the stamp was cut off around the microstructures together with the PDMS layer, and the layer was carefully peeled off, turned upside down, and laid down on a glass cover slip. The thickness of the resulting PDMS layer, which was measured to be $101 \pm 2 \mu\text{m}$ by profilometer, was large enough to allow easy handling, i.e. to peel off and to lay down on a glass cover slip, and small enough to allow high resolution microscopy observation of cells on the layer. Two different molecules were tested for silanization of the stamp surface, TMCS and (tridecafluoro-1,1,2,2-tetrahydrooctyl)trichlorosilane. For stamps with microstructures carrying the smallest feature (5 μm), only the latter molecule, which exposes on the surface fluorocarbon chain, allowed smooth and defect-less removal of the molded PDMS layer. This molecule was thus used for all the experiments.

Micro-grooves (μGro). μGro consists in 5 μm -high microstructures with uniform coating of PLL-FITC. To obtain a thin layer of PDMS with the microstructures, non-reticulated, degassed PDMS mixture was spread on a mold by spin-coating at 800 rpm, the same speed as for iMP, for 30 s prior to incubation at 66 $^{\circ}\text{C}$ overnight, peeled off, and laid down on a glass cover slip (Fig. 2, bottom). The mold was fabricated by following the same procedure as for μCP and iMP, but with a photomask of inverted contrast. PLL-FITC coating was achieved after microfluidic chip assembly (see below).

Geometry of micro-patterns on the photomask

All the micro-patterns on the photomasks used for the mold preparation of μCP , iMP and μGro substrates contained a $900 \mu\text{m} \times 8 \text{mm}$ rectangle, which resulted in a cell seeding area homogeneously coated with PLL-FITC on the final substrates. From one of the long sides of the rectangle, 1 mm-long, straight, and parallel lines with different width ($w = 5, 10, 20, 40, \text{ and } 80 \mu\text{m}$) and intervals ($d = 5, 10, 20, 40, \text{ and } 80 \mu\text{m}$), which served as axon guiding tracks, projected out perpendicularly to the rectangle. The length of the rectangle was divided in several 700 μm -long blocks corresponding to a given pair of the (w, d) values (see Fig. 3, bottom right). We used a chromium mask for the micro-patterns containing structures characterized by d or w equal to 5 μm , and a transparency mask with the resolution of 50800 dpi for the others. Importantly, for μCP and iMP, the background outside the micro-patterns was filled with 50 μm -diameter disks arranged at 100 μm pitch. These disks resulted in pillars on the PDMS stamps, which prevented the stamps from collapsing on the substrate.

Microfluidic chip preparation

Microfluidic chip with a removable partition. On top of the micro-patterned substrate described in the previous sections, a PDMS block with a microfluidic channel equipped with a removable partition, i.e. a microfilament, was bonded to build up the complete microfluidic chip (Fig. 3). This novel, yet simple microfluidic system allowed a temporary division of the microfluidic channel into two independent compartments. To fabricate the microfluidic channel, an 80 μm -high SU-8 mold was prepared by spin-coating SU-8 2050 (MicroChem) on a 2 inch quartz disk (Neyco) at 2000 rpm for 30 s followed by photolithography and silanization with TMCS, and PDMS with a thickness about 6 mm was reticulated on the mold and cut off. The microfluidic channel involves a 2 mm-wide and 9 mm-long straight part, which splits at both ends into a thin 70 μm -wide and 8 mm-long middle straight channel (guiding duct) surrounded by two 900 μm -wide and 2.5 mm-long channels connected to reservoirs (Fig. 3, top left). These reservoirs were made by punching through the PDMS block using a 4 mm-diameter biopsy punch (Kai Industries). As a removable partition, we employed a fishing line (MORRIS) with a diameter of 74 μm in nylon (VARIVAS Super tippet Master spec 9X, for μCP) or 86 μm in fluorocarbon (VARIVAS Fluoro carbon Super tippet 9X, for μGro and iMP). The PDMS block was pierced twice by the line at the ends of the guiding ducts using a sewing needle, and the line was inserted into the channel using the ducts as guiding structures (Fig. 3, middle left). After being tensed, the line was further embedded in the guiding ducts using tweezers, followed by chip assembly (see below). This cylindrical wall separates a flat, uniformly coated area (i.e. cell seeding area) from the micro-patterned part (i.e. axon guidance area) of the bottom substrate in the channel, as shown in Fig. 3, top right. The partition can be easily removed by pulling the line, which is freely sliding in the PDMS device along the ducts.

Chip assembly. Two 4 cm-long pieces of Scotch removable tape were folded in half along the main axis with their non-adhesive side out, and used as spacers. The measured thickness of the spacer was 100 μm . The PDMS block with the microfluidic channel and the fishing line was cleaned with ethanol, dried with pressurized air flow, plasma treated, and used immediately. For iMP, the spacers were inserted between the substrate and the PDMS block to avoid their contact, and the microstructure on the substrate and the microfluidic channel were aligned under microscope. The fishing line was aligned at the beginning of the tracks (see Fig. 3, top right). Once aligned, the PDMS block was pushed down to be in contact with the substrate, and the spacers were removed to complete the bonding of the PDMS block and the substrate.³² Immediately after bonding, DPBS was introduced in the microfluidic channel with filling each reservoir with 40 μL of DPBS by micropipette. Air bubbles trapped in the channel were squeezed out, if necessary. For μGro , the substrate was plasma treated prior to use, and the same process as for iMP was applied with replacing DPBS with 0.01 mg/mL PLL-FITC in DPBS. For μCP , the alignment was achieved following the paper template of the microfluidic channel fixed under the substrate, and the same process as for iMP was applied. To avoid evaporation, all the chips were enclosed in a plastic box containing water saturated with ethylenediaminetetraacetic acid and kept in a humidified CO_2 incubator overnight at 37 $^\circ\text{C}$. Finally, the microfluidic channel was rinsed twice with DPBS and incubated with complete culture medium: DMEM GlutaMAX high glucose (DMEM) supplemented with 10 % fetal bovine serum (FBS), 2 % B27, 1 % N2 (v/v), 100 U/mL Penicillin-Streptomycin (all from Life Technologies), for about 6 h in the CO_2 incubator before cell seeding.

Density of PLL coating according to the different types of substrates

As described above, PLL-FITC was directly incubated on μGro samples, while it was transferred between PDMS surfaces once and twice for μCP and iMP, respectively, during substrate preparation. These different protocols yielded different coating density on the final chip, as displayed in ESI Fig. S1a. PLL-FITC density was the highest on μCP samples, and the lowest for the iMP technique with a reduction by a factor of 5 compared to μCP . The spatial distribution of PLL-FITC was also assessed for the three different techniques, showing that the finer, i.e. 5 μm wide micro-pattern separated by 5 μm interval was successfully obtained (ESI Fig. S1b) and that the PLL-FITC layer was uniform in all techniques (ESI Fig. S1c-e).

Cell culture

The study was carried out in accordance with European Community guidelines on the care and use of laboratory animals: 86/609/EEC. Cortices were dissected under microscope from E17 embryos of Oncins France 1 mouse (Charles River) in Gey's balanced salt solution supplemented with 0.1 % glucose (w/v). After rinsing, the cortices were incubated for 15 min at 37 $^\circ\text{C}$ with 20 U/mL papain and 0.6 mg/mL L-cysteine in DMEM supplemented with 100 U/mL Penicillin-Streptomycin. After deactivating papain by adding one-tenth volume of FBS, followed by 5 min incubation at room temperature and rinsing with DMEM, cortical cells were mechanically dissociated by pipetting in DMEM containing 40 U/mL deoxyribonuclease I. The cells were rinsed twice with DMEM and suspended in complete culture medium (see above) at 3×10^7 cells/mL. After removal of the medium from the reservoirs of the microfluidic chip, 2 μL each of cell suspension and complete medium were introduced simultaneously in the chip from one reservoir on the cell-seeding-area side and from the next one on the micro-patterned side, respectively. The chip was incubated in the CO_2 incubator at 37 $^\circ\text{C}$ with 5 % CO_2 for 10 min to make cells sediment and attach on the substrate, and then each reservoir of the chip was filled with 40 μL of complete culture medium. The chip was enclosed in the plastic box described above to avoid evaporation, and placed in the CO_2 incubator. One hour after cell seeding, one extremity of the fishing line was cut off at

the exit from the PDMS block, and the line was removed from the chip by pulling the other extremity with tweezers. Cells were cultured for 6 or 11 days in the incubator, as previously described.²⁶ For 11 days culture, complete medium was renewed on the 6th day.

Immunofluorescence staining and imaging

After removal of culture medium from the reservoirs of the chip, 4 % paraformaldehyde (Electron Microscopy Sciences, v/v) in DPBS supplemented with 4 % sucrose (w/v) was introduced in the channel for cell fixation, using a slight gradient of hydrostatic pressure in the longitudinal direction. For that purpose, 30 μL of solution was added to each upstream reservoir, and 20 μL to the downstream ones, using an automatic pipette (Finnpipette, Thermo Scientific). The chip was incubated for 10 min at room temperature and rinsed with DPBS. DPBS containing 1 % bovine serum albumin (w/v) and 0.1 % Triton-X (Life Technologies, v/v) was introduced in the chip with a hydrostatic pressure gradient as described above, and incubated for 45 min at room temperature. After rinsing with DPBS, a primary antibody solution containing anti-microtubule-associated protein 2 (MAP2) antibody (EMD Millipore, rabbit polyclonal, 1/300 volume fraction) and anti-Tau antibody (EMD Millipore, mouse monoclonal, 1/300) in DPBS was introduced in the channel with a gradient (20 μL each to the upstream reservoirs and 15 μL each to the downstream ones) and incubated overnight at 4 °C. After rinsing with DPBS, a secondary antibody solution containing Alexa Fluor 546-conjugated anti-rabbit IgG antibody (Life Technologies, goat polyclonal, 1/200), Alexa Fluor 647-conjugated anti-mouse IgG antibody (Life Technologies, goat polyclonal, 1/200), and 1 $\mu\text{g}/\text{mL}$ Hoechst (Life Technologies) in DPBS was introduced in the same manner as the primary antibody solution, and incubated for 2 h at room temperature. After rinsing twice with DPBS, the sample was observed by fluorescence microscopy (Eclipse Ti, Nikon) equipped with a CCD camera (CoolSNAP HQ2, Photometrics) through a 4 \times objective or by confocal microscopy (LSM 510 Meta, Zeiss) through 40 \times and 63 \times objectives, with sealing the reservoirs with 10 μL each of mineral oil. Images were analyzed with ImageJ software.

Results

All types of substrates yielded similar maximal axonal lengths integrated over 6 DIV (ESI Fig.S2a, b), indicating that the growth rate is not compromised by the significant reduction in the density of PLL coating associated to iMP (ESI Fig. S1a). Similarly, we did not observe any difference in the density of soma attached in the seeding areas of the different types of substrates (ESI Fig. S2c). This suggests that even in iMP, the lower density of PLL is not detrimental to cell soma adhesion. However, the extent and the characteristics of the confinement differ from one constraint to the other, as revealed by different behaviors occurring (i) between the tracks, (ii) at the track ends and (iii) along the tracks. Qualitative and quantitative analysis of these differences in axonal confinement are provided below.

Frequency of barrier crossing

The most dramatic difference between techniques is revealed quantitatively at 11 DIV (see figure 4a for immunofluorescence images of neurons at 11 DIV in the case of 5 μm -wide tracks with 5 μm intervals) by the variation in the number of barrier crossing events per unit length, n_c , depending on the interval of stripes, d . To evaluate n_c , the length of the part of a track filled with axons, i.e. L , was first measured. Then, for the simplicity of the measurement, the number of axons escaping from the track on both sides as 'branches', n_b , was counted regardless of whether or not the axons were reaching the next track (see the scheme of this measurement in Fig. 4b). Parts of a track where crossing events were affected by extrinsic factors or finite-size effects were excluded from the measurement. This corresponds for instance to axons reflected at the end of a track (seen in μCP and μGro), or rare cases in which axons grew on a barrier directly from a soma positioned on the barrier (seen in μCP and μGro), or perturbation of axon growth by glial cells on the barrier (seen in μGro). The average number of crossings per length was finally defined as $n_c = \Sigma n_b / (2\Sigma L)$, since the number of crossings could be roughly estimated as $n_b/2$ (e.g. an axon totally crossing the barrier,




Figure 4. Efficiency of axon confinement in the tracks. (a) Immunofluorescence images of neurons at 11 DIV on different substrates with 5 μ m-wide guiding tracks with 5 μ m intervals. Axons (green), dendrites (red), and nuclei (blue) are stained by using anti-Tau, anti-MAP2 antibodies, and Hoechst, respectively. Scale bars, 100 μ m. (b) Scheme of the counting of the number of barrier crossing events per unit length, n_c (left). (c) n_c is displayed against the track interval, d , for the different techniques. iMP shows a remarkably high efficiency of axon confinement. For $d = 5, 10, 20, 40, 80$ μ m, number of samples are: $n = 104, 319, 268, 153, 92$ (μ CP); $n = 463, 337, 259, 153, 119$ (μ Gro); $n = 533, 476, 266, 167, 120$ (iMP), respectively. Error bars represent standard deviations.

corresponding to $n_c = 1$, has $n_b = 1$ on one track and $n_b = 1$ on the next track). The summation was taken over a large number of samples from the same dissection so that Σn_b became non-zero.

Figure 4c shows the values of n_c , averaged among 2 to 5 sample groups from independent dissections, depending on d for different substrates. The values of n_c for iMP are 1 to 2 orders of magnitude smaller than those for the other techniques, proving that iMP has by far the highest efficiency of axon confinement among these substrates. In the case of μ CP, n_c is high relative to any other situations when the track interval, i.e. barrier length, d is 5 μ m. It is still significant for $d = 10$ μ m, and then drops for higher d values. We associate this poor efficiency at low spacing with the capability of filipodia to explore 2D areas even in the absence of PLL coating. This assumption is consistent with an earlier work that estimated a filipodia length around 5 μ m.³³ For μ Gro, n_c is higher than for iMP for all values of d , and becomes larger than that for μ CP for $d \geq 10$ μ m. This limited confinement effect indicates that the 5 μ m-high obstacle provided by the border is easily crossed by axons when the whole chip surface is adhesive, in agreement with previous results.³⁴ The increase of n_c with the increase of d in μ Gro can be simply explained by the

fact that the larger is d , the higher is the probability that an axon branches before reaching the next track (as shown in the scheme of Fig. 4b). This last analysis also provides a measure of the characteristic length between axonal branching points, about 40 μm , also in agreement with previous observations.³⁵

In the case of iMP, n_c is remarkably small ($n_c = 0.009, 0.018, 0.022, 0.033, 0.015$ par mm for $d = 5, 10, 20, 40, 80 \mu\text{m}$, respectively) and axons very rarely escape from tracks. Note that iMP at $d = 5 \mu\text{m}$ gives n_c values one order of magnitude smaller than those of μCP at $d = 20 \mu\text{m}$, whereas the cross-sectional contour length of the 5 μm -high barrier at $d = 5 \mu\text{m}$, i.e. 15 μm (representing the transverse length axons must cross in order to jump from one groove to the other), is smaller than 20 μm . This provides a direct evidence that a non-coated topographical barrier strongly improves the confinement efficiency, and is more efficient in that respect than increasing the length of a flat, non-coated interval between tracks. Conversely, it has been shown that when only topographical barriers with cell adhesive surfaces, like as in μGro condition, are applied, 25 μm -high walls are necessary to achieve a high efficiency confinement in tracks.^{6, 34} By combining both chemical and topographical constraints, we succeeded to improve drastically the guidance and confinement efficiency.

Fate of axons at the track ends

Typical behaviors associated to each type of constraints are observed when axons reach the track ends, as shown in Fig. 5a. In the case of μCP , axons tend to keep on growing by turning at right angle and jumping from one neighboring tracks to the other. This phenomenon is associated to a relatively higher probability of crossing of the non-adhesive areas between PLL-coated stripes for μCP , as described in the previous section. In the case of μGro , axons are allowed to continue their growth out from the tracks due to the homogenous PLL-FITC coating. Interestingly, axons show a directional persistency outside the grooves. In contrast, in the condition of iMP, axons stop at the walls located at the end of the tracks. This is an additional evidence of the strong confinement effect of iMP, since even axons impinging onto the wall in a head-on trajectory are unable to escape the groove, whereas some are able in μGro .




Fig. 5. Axon behaviors at the end and within tracks. (a) Immunofluorescence of axons (green), stained by using anti-Tau antibody, at the end of the tracks on different techniques. (b) Immunofluorescence of axons (anti-Tau, green) guided on 80 μm -wide tracks in the direction from the bottom to the top of the images. (c) Normalized profiles of the integrated fluorescence intensity of axons across the tracks. Thin lines in each graph represent the profiles of 10 different samples and the thick line represents their average. (d) A projection of cross-sectional images of PLL-FITC fluorescence (magenta) and immunofluorescence of axons (anti-Tau, green) on μGro 20 μm -wide tracks at 6 DIV, obtained from a three dimensional reconstruction of a 18.5 μm -thick stack of 143 $\mu\text{m} \times 143 \mu\text{m}$ confocal microscopy images. The fluorescence of anti-Tau, reflecting the presence of axons, is accumulated at the corners of grooves (arrowheads). Scale bars, 50 μm .

Axonal behavior inside the tracks

Axons adopt different types of trajectory according to the track widths, and most of all according to the nature of the substrate. We observed that when axons were guided on a track wider than 20 μm , i.e. a track of 40 μm or 80 μm width, the track surface was in general not totally covered by axons even at 11 DIV (Fig. 5b). A peak of the fluorescence intensity of Tau, i.e. a preferential positioning of axons along the edges of the tracks can be evidenced for the three types of guiding tracks. This phenomenon is however particularly pronounced for μGro (Fig. 5c). Confocal microscopy observation of PLL-FITC and axon immunofluorescence was performed on a fixed culture at 6 DIV on a μGro substrate, when axons are not highly bundled. The observation (Fig. 5d) revealed that axons preferably grow at the edges of the grooves on the bottom surface along the sidewalls, as indicated with arrowheads, and not on the groove walls, in agreement with a previous report.⁶

Axon guidance from small independent populations containing several neurons by using iMP

We have shown above the interest of iMP for axonal confinement. We have further explored the potentiality of this hybrid constraint with the aim of compartmentalizing soma, thus dividing the seeding area into an array of small circular microchambers (Fig. 6). Different diameters, from 15 to 100 μm were explored (data not shown). With the concentration of the cellular suspension used in this study, a diameter of 40 μm yielded a systematic filling by one cell at least, with a maximum of about 10. In this particular example, axons were guided by 5 μm -wide tracks separated by a distance of 80 μm connected to these small circular chamber (Fig. 6). Similarly to what was done in larger chambers, a microfilament was placed at a distance of about 100 μm from the circles as indicated with a dotted line in Fig. 6a, and removed 1 h after cell seeding. The experiment demonstrates that iMP technique allows to build a regular array of fully independent and organized neuronal compartments.




Figure 6 Axon guidance from small independent populations of neurons by using iMP technique. (a) An immunofluorescence image of axons (anti-Tau, green), dendrites (anti-MAP2, red), and nuclei (Hoechst, blue) at 6 DIV (left) and a transmission image (right). Neurons were seeded and trapped in 40 μm -diameter circles by using the same system as the other experiments. A fishing line was placed at the distance of about 100 μm from the circles as indicated with a dashed line, and removed 1 hour after cell seeding. Axons were guided by 5 μm -wide tracks connected to the circles. (b) Transmission and immunofluorescence images corresponding to the area surrounded by rectangles in (a). Scale bars, 100 μm .

Discussion

Among the three methods developed to guide axon growth in an open configuration tested here, the new approach by in-mold patterning presents strikingly superior performances, with typically 10 to 100 times less track escapes than uniformly coated microgrooves or micro-contact printing. The underlying strategy, of adding chemical and physical guiding cues, is thus a fruitful one. Interestingly, this difference is obtained without spectacular changes in other features of axon growth: the growth speed, and the proportion of axons reaching the end of the track after 6 days, are relatively similar, with a small advantage to μ Gro that is further discussed in ESI text S1.

Besides its superior guidance efficiency, iMP has several additional advantages over μ Gro and μ CP. During imaging, in contrast with μ CP, the position of adhesive patterns can be identified without the need of fluorescently labelled adhesive molecules that bring in a series of disadvantages: labelling may affect adhesion properties or yield some toxicity, the presence of a fluorescent background reduces the choice of fluorophores for cell imaging, and reduces signal-to-noise ratio. Essentially, all uses of guiding patterns developed in the context of earlier guiding cues can be transposed to iMP with improved performances. For instance, neurite selection by pattern width and shape¹⁶ are also transposable to iMP, with a stronger and longer lived confinement effect. Additionally, axons or cell bodies at the bottom of the microstructures are protected from shear stress in the microfluidic channel, if the microstructures have a significantly high aspect ratio (typically >1). Cells can be sensitive to shear stress, which can modify their phenotype or viability, and neurons are in this respect particularly fragile.³⁶

Regarding neuron guidance, all the patterns developed within the microchannel paradigm,²⁴⁻²⁷ such as “axon diodes”,²⁶ can be transposed to iMP, retaining the advantages of the above for neurite guidance, selectivity and/or polarization, while remaining in an open configuration.

iMP also opens new possibility for the easy construction of complex, multilayer networks, out of reach of current patterning or microchannel devices. First, multilayer systems with embedded chemical patterns can be prepared much more easily than previously: the micro-pattern is visible in a simple microscope or binocular, without fluorescence. This property facilitates, during the assembly of a multilayer chip, the alignment between the substrate and features such as microfluidic channels, chambers, electrodes, etc. Since the bonding of the layers of the microfluidic chip does not occur on a plane occupied by axons, the chip can be assembled without plasma treatment and without risk of disjoining of the two parts during growth of axons, as observed in microchannel systems in the absence of permanent bonding. The chip can thus be opened after axon growth without disturbing the neuronal array.²⁸ Such an open configuration in turn allows direct access or addressing by chemical or physical means, such as mechanical or electrical sensing, microelectrodes, micropipettes, microlever, patch-clamp, and the like.

“Hybrid” devices, combining guidance by iMP and fluidic insulation by microchannels, can be prepared easily, for instance by partially sealing the guiding tracks with a PDMS block with macro-chambers. This could be interesting for combining fluidic separation and axonal communication between macro-chambers, while retaining guidance in the macro-chambers. This can allow, for instance, following identical axons crossing macro-chambers containing different components, improving significantly the conventional on-chip neuron culture platforms.²⁴⁻²⁷ It is also possible to perform μ CP on an iMP substrate, thus opening the route to multilayers guiding structures with overpasses and underpasses.

Finally, as demonstrated in Fig. 6, iMP can also be used for the positioning of cell soma within independent micro-chambers, with a very high selectivity. Combined with a gradient of active molecules in solution (see ⁹ for an example of a combination of topographic constraints with a gradient of soluble cues), such an array of independent neuronal compartments would allow to probe independent cell responses to different concentrations of e.g. a drug on a single chip.

Regarding future possible developments, one should also remark that iMP as prepared here is patterned onto a thin layer of PDMS. Thin PDMS films without microstructures have been used in the past in stretching assays, to study mechanobiology of neurons³⁷ or axonal injury.³⁸ iMP structures might therefore increase the power of such studies, e.g. by imposing a predefined angle between the direction of the applied mechanical force and the direction of the axons in order to mimic some pathological situations.³⁹ Moreover, iMP is not limited to the use of cell adhesive molecules such as PLL. For instance, carbon nanotube electrodes can be integrated in a PDMS layer by iMP technique.⁴⁰ Deformable micro-electrode arrays on a PDMS layer have also been proposed.⁴¹⁻⁴³ By integrating such micro-electrodes into axon guiding tracks with high confinement efficiency, iMP could thus improve the specificity and efficiency of addressing in microelectrode array (MEA) devices.

Conclusions

In this work, we have studied the effect on the growth of axons of three different types of cellular confinement, i.e. topographical (micro-grooves), chemical (micro-contact printing), and a hybrid method combining topographical and chemical cues, called in-mold patterning. In this approach, an adhesive layer, here PLL, is deposited at the bottom of microstructures only, leaving the walls of the structures and the upper surface non-adhesive. The potential of these different approaches for guiding and confining axons was studied using linear patterns with sizes and spacing ranging from 5 to 80 μm . Overall, in-mold patterning, applied here for the first time to neurons, provides a confinement in guiding structures 10 to 100 times higher than the two other techniques, and is also able to accurately position cell bodies in microwells. This technology was advantageously combined for initial seeding with a transient partitioning technique allowing the confinement of soma and the extension of axons in separate areas of a microfluidic chips.

On the microfabrication side, in-mold patterning does not involve more complex microfabrication technologies than already needed for micro-grooves, micro-contact printing, or enclosed microchannels, or significantly longer manufacturing time as compared to μCP . The iMP technique will thus be available to numerous laboratories. It simplifies the alignment of multilayers structures with chemical micro-patterns, thus opening the route to the easy fabrication of three-dimensional arrays with overpasses and underpasses. Overall, it offers the same potential for cell positioning, confinement, and guidance than systems based on enclosed microchannels, but in an open environment giving access to a number of new experimental approaches, such as microelectrodes, micropipettes, local sensors or actuators, or chemical addressing by movable fluidic boundaries. We thus believe that it will constitute a new and important component of the growing toolbox for cell positioning, manipulation and addressing, and more generally for experimental neurosciences.

Acknowledgements

We thank Céline Braïni and the Nanofab platform in Institut Néel (Grenoble) for their help with our experiments. We thank Ivan Ferrante, Andreas Offenhäuser, Bernhard Wolfrum, Bastian Haberkorn, Philippe Tixador, Bérangeère Deleglise, Anselme Perrier, and Simona Gribaudo for fruitful discussions and their technical advices. This work was performed in part in the UMR 168 microfabrication cleanroom and in the BioImaging Cell and Tissue Core Facility of the Institut Curie (PICT-IBiSA). This work was supported by ANR project Neuroscreen, grants from the European Commission: MicroDEG, ERA-NET Neuron JTC2012 "Novel Methods", <http://www.neuron-eranet.eu/en/317.php>, and European research Council Advanced Grant N°:321107 "CellIO". BV acknowledges support from DGA.

Notes and references

1. C. Tomba and C. Villard, *Microelectron. Eng.*, 2015, **132**, 176-191.
2. Y. Xia and G. M. Whitesides, *Angew. Chem. Int. Ed.*, 1998, **37**, 550-575.
3. D. W. Branch, J. M. Corey, J. A. Weyhenmeyer, G. J. Brewer and B. C. Wheeler, *Med. Biol. Eng. Comput.*, 1998, **36**, 135-141.
4. A. Offenhäuser, S. Böcker-Meffert, T. Decker, R. Helpenstein, P. Gasteier, J. Groll, M. Möller, A. Reska, S. Schäfer, P. Schulte and A. Vogt-Eisele, *Soft Matter*, 2007, **3**, 290-298.
5. T. L. Lien, J. Ban, M. Tormen, E. Migliorini, G. Grecni, A. Pozzato and V. Torre, *PLoS One*, 2013, **8**, e73966.
6. A. Beduer, I. Gonzales-Calvo, C. Vieu, I. Loubinoux and L. Vaysse, *Macromol. Biosci.*, 2013, **13**, 1546-1555.
7. J. Park, S. Kim, S. I. Park, Y. Choe, J. Li and A. Han, *J. Neurosci. Methods*, 2014, **221**, 166-174.
8. N. M. Dowell-Mesfin, M. A. Abdul-Karim, A. M. Turner, S. Schanz, H. G. Craighead, B. Roysam, J. N. Turner and W. Shain, *J. Neural Eng.*, 2004, **1**, 78-90.
9. A. Kundu, L. Micholt, S. Friedrich, D. R. Rand, C. Bartic, D. Braeken and A. Levchenko, *Lab Chip*, 2013, **13**, 3070-3081.
10. M. J. Jang and Y. Nam, *J. Neural Eng.*, 2012, **9**, 046019.
11. E. Marconi, T. Nieus, A. Maccione, P. Valente, A. Simi, M. Messa, S. Dante, P. Baldelli, L. Berdondini and F. Benfenati, *PLoS One*, 2012, **7**, e34648.
12. J. Zhang, S. Venkataramani, H. Xu, Y. K. Song, H. K. Song, G. T. Palmore, J. Fallon and A. V. Nurmikko, *Biomaterials*, 2006, **27**, 5734-5739.
13. S. B. N. Biancardo, H. J. Pranov and N. B. Larsen, *Adv. Mater.*, 2008, **20**, 1825-1829.
14. M. Hakanson, S. Kobel, M. P. Lutolf, M. Textor, E. Cukierman and M. Charnley, *PLoS One*, 2012, **7**, e40141.
15. S. Roth, G. Bugnicourt, M. Bisbal, S. Gory-Faure, J. Brocard and C. Villard, *Small*, 2012, **8**, 671-675.
16. C. Tomba, C. Braini, B. Wu, N. S. Gov and C. Villard, *Soft Matter*, 2014, **10**, 2381-2387.
17. R. Fricke, P. D. Zentis, L. T. Rajappa, B. Hofmann, M. Banzet, A. Offenhäuser and S. H. Meffert, *Biomaterials*, 2011, **32**, 2070-2076.
18. H. Hardelauf, J. Sisnaiske, A. A. Taghipour-Anvari, P. Jacob, E. Drabiniok, U. Marggraf, J. P. Frimat, J. G. Hengstler, A. Neyer, C. van Thriel and J. West, *Lab Chip*, 2011, **11**, 2763-2771.

19. J. P. Frimat, J. Sisaïsk, S. Subbiah, H. Menne, P. Godoy, P. Lampen, M. Leist, J. Franzke, J. G. Hengstler, C. van Thriel and J. West, *Lab Chip*, 2010, **10**, 701-709.
20. I. Suzuki, Y. Sugio, Y. Jimbo and K. Yasuda, *Lab Chip*, 2005, **5**, 241-247.
21. W. Li, Z. Xu, J. Huang, X. Lin, R. Luo, C.-H. Chen and P. Shi, *Sci. Rep.*, 2014, **4**, 4784.
22. A. Kunze, M. Giugliano, A. Valero and P. Renaud, *Biomaterials*, 2011, **32**, 2088-2098.
23. L. J. Millet and M. U. Gillette, *Trends Neurosci.*, 2012, **35**, 752-761.
24. A. M. Taylor, M. Blurton-Jones, S. W. Rhee, D. H. Cribbs, C. W. Cotman and N. L. Jeon, *Nat. Methods*, 2005, **2**, 599-605.
25. J. W. Park, B. Vahidi, A. M. Taylor, S. W. Rhee and N. L. Jeon, *Nat. Protoc.*, 2006, **1**, 2128-2136.
26. J. M. Peyrin, B. Deleglise, L. Saias, M. Vignes, P. Gougis, S. Magnifico, S. Betuing, M. Pietri, J. Caboche, P. Vanhoutte, J. L. Viovy and B. Brugg, *Lab Chip*, 2011, **11**, 3663-3673.
27. N. D. Dinh, Y. Y. Chiang, H. Hardelauf, J. Baumann, E. Jackson, S. Waide, J. Sisaïsk, J. P. Frimat, C. van Thriel, D. Janasek, J. M. Peyrin and J. West, *Lab Chip*, 2013, **13**, 1402-1412.
28. A. Yamada, B. Venzac, I. Pereiro, S. Coscoy, M. Verhulsel, M. C. Parrini, J.-L. Viovy and S. Descroix, Submitted.
29. A. C. v. Philipsborn, S. Lang, A. Bernard, J. Loeschinger, C. David, D. Lehnert, M. Bastmeyer and F. Bonhoeffer, *Nat. Protoc.*, 2006, **1**, 1322-1328.
30. W. R. Kim, M. J. Jang, S. Joo, W. Sun and Y. Nam, *Lab Chip*, 2014, **14**, 799-805.
31. J. C. Chang, G. J. Brewer and B. C. Wheeler, *Biomaterials*, 2003, **24**, 2863-2870.
32. T. Hamkins-Indik, S. Lam, M. E. Dueck and L. P. Lee, *Lab Chip*, Chips and Tips 17 Aug 2011.
33. A. Valerio, V. Ghisi, M. Dossena, C. Tonello, A. Giordano, A. Frontini, M. Ferrario, M. Pizzi, P. Spano, M. O. Carruba and E. Nisoli, *J. Biol. Chem.*, 2006, **281**, 12950-12958.
34. N. Li and A. Folch, *Exp. Cell Res.*, 2005, **311**, 307-316.
35. A. R. Kriegstein and M. A. Dichter, *J. Neurosci.*, 1983, **3**, 1634-1647.
36. M. Morel, V. Shynkar, J. C. Galas, I. Dupin, C. Bouzigues, V. Studer and M. Dahan, *Biophys. J.*, 2012, **103**, 1648-1656.
37. J. Chetta, C. Kye and S. B. Shah, *Cytoskeleton*, 2010, **67**, 650-665.
38. B. J. Pfister, T. P. Weihs, M. Betenbaugh and G. Bao, *Annal. Biomed. Eng.*, 2003, **31**, 589-598.
39. R. J. Cloots, J. A. van Dommelen, T. Nyberg, S. Kleiven and M. G. Geers, *Biomech. Model. Mechanobiol.*, 2011, **10**, 413-422.
40. Y. Hanein, *Phys. Status Solidi B*, 2010, **247**, 2635-2640.
41. L. Guo, G. S. Guvanasen, X. Liu, C. Tuthill, T. R. Nichols and S. P. DeWeerth, *IEEE Trans. Biomed. Circuits Syst.*, 2013, **7**, 1-10.
42. L. Bareket-Keren and Y. Hanein, *Front. Neural Circuits*, 2013, **6**, 122.
43. S. P. Lacour, S. Benmerah, E. Tarte, J. FitzGerald, J. Serra, S. McMahon, J. Fawcett, O. Graudejus, Z. Yu and B. Morrison, 3rd, *Med. Biol. Eng. Comput.*, 2010, **48**, 945-954.