



**HAL**  
open science

## Joint Position Paper of the Working Group of Pacing and Electrophysiology of the French Society of Cardiology and the French Society of Diagnostic and Interventional Cardiac and Vascular Imaging on magnetic resonance imaging in patients with cardiac electronic implantable devices

Estelle Gandjbakhch, Jean-Nicolas Dacher, Jerome Taieb, Michel Chauvin, Frederic Anselme, Axel Bartoli, Louis Boyer, Lucie Cassagnes, Hubert Cochet, Pascal Defaye, et al.

### ► To cite this version:

Estelle Gandjbakhch, Jean-Nicolas Dacher, Jerome Taieb, Michel Chauvin, Frederic Anselme, et al.. Joint Position Paper of the Working Group of Pacing and Electrophysiology of the French Society of Cardiology and the French Society of Diagnostic and Interventional Cardiac and Vascular Imaging on magnetic resonance imaging in patients with cardiac electronic implantable devices. Archives of cardiovascular diseases, 2020, 113 (6-7), pp.473-484. 10.1016/j.acvd.2020.03.015 . hal-02949307

**HAL Id: hal-02949307**

<https://hal.sorbonne-universite.fr/hal-02949307v1>

Submitted on 22 Aug 2022

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

**Joint Position Paper of the Working Group of Pacing and Electrophysiology of the French Society of Cardiology and the French Society of Diagnostic and Interventional Cardiac and Vascular Imaging on magnetic resonance imaging in patients with cardiac electronic implantable devices**

*Position conjointe du groupe de rythmologie et de stimulation cardiaque de la Société Française de Cardiologie et de la Société Française d'Imagerie Cardiaque et Vasculaire Diagnostique et Interventionnelle sur la pratique des examens par résonance magnétique chez les patients porteurs de dispositifs cardiaques électroniques implantables*

**Abbreviated title:** MRI and cardiac electronic implantable devices

**Estelle Gandjbakhch<sup>a,\*</sup>, Jean Nicolas Dacher<sup>b</sup>, Jerome Taieb<sup>c</sup>, Michel Chauvin<sup>d</sup>, Frederic Anselme<sup>e</sup>, Axel Bartoli<sup>f</sup>, Louis Boyer<sup>g</sup>, Lucie Cassagnes<sup>g</sup>, Hubert Cochet<sup>h</sup>, Pascal Defaye<sup>i</sup>, Jean-Claude Deharo<sup>j</sup>, Benjamin Dubourg<sup>b</sup>, Laurent Fauchier<sup>k</sup>, Daniel Gras<sup>l</sup>, Didier Klug<sup>m</sup>, Gabriel Laurent<sup>n</sup>, Jacques Mansourati<sup>o</sup>, Eloi Marijon<sup>p</sup>, Philippe Maury<sup>q</sup>, Olivier Piot<sup>r</sup>, François Pontana<sup>s</sup>, Frederic Sacher<sup>t</sup>, Nicolas Sadoul<sup>u</sup>, Alexis Jacquier<sup>f</sup>, Serge Boveda<sup>v</sup>, for the Working Group of Pacing and Electrophysiology of the French Society of Cardiology and the French Society of Diagnostic and Interventional Cardiac and Vascular Imaging**

<sup>a</sup> Sorbonne University, Heart Institute, La Pitié-Salpêtrière University Hospital, AP-HP, 75013 Paris, France

<sup>b</sup> Cardiac Imaging Unit, Department of Radiology, Rouen University Hospital, 76031 Rouen, France

<sup>c</sup> Department of Cardiology, Aix-en-Provence Hospital, 13100 Aix-en-Provence, France

<sup>d</sup> Department of Cardiology, University Hospital, 67000 Strasbourg, France

<sup>e</sup> Department of Cardiology, Rouen University Hospital, 76000 Rouen, France

<sup>f</sup> Department of Radiology, La Timone 2 University Hospital, Aix-Marseille University, CNRS, CRMBM, CEMEREM, 13005 Marseille, France

<sup>g</sup> Department of Radiology, University Hospital Gabriel Montpied, 63000 Clermont-Ferrand, France

<sup>h</sup> *Department of Cardiovascular Imaging, Haut-Lévêque Cardiological Hospital, CHU de Bordeaux, IHU LIRYC, Université de Bordeaux-Inserm, 33600 Pessac, France*

<sup>i</sup> *Department of Cardiology, Grenoble University Hospital, 38700 La Tronche, France*

<sup>j</sup> *Department of Cardiology, La Timone University Hospital, 13005 Marseille, France*

<sup>k</sup> *Department of Cardiology, Tours University Hospital, 37000 Tours, France*

<sup>l</sup> *Nouvelles Cliniques Nantaises, 44200 Nantes, France*

<sup>m</sup> *Department of Cardiology, Lille University Hospital, 59000 Lille, France*

<sup>n</sup> *Department of Cardiology, Dijon University Hospital, 21000 Dijon, France*

<sup>o</sup> *Department of Cardiology, Brest University Hospital, 29200 Brest, France*

<sup>p</sup> *Department of Cardiology, Georges Pompidou European Hospital, AP-HP, 75015 Paris, France*

<sup>q</sup> *Department of Cardiology, Rangueil University Hospital, 31059 Toulouse; Unité Inserm U1048, 31432 Toulouse, France*

<sup>r</sup> *Centre Cardiologique du Nord, 93200 Saint-Denis, France*

<sup>s</sup> *Department of Cardiovascular Radiology, Institut Cœur-Poumon, INSERM UMR 1011, Lille University Hospital, 59000 Lille, France*

<sup>t</sup> *Department of Cardiology, Haut-Lévêque Cardiological Hospital, CHU de Bordeaux, IHU LIRYC, Université de Bordeaux-Inserm, 33600 Pessac, France*

<sup>u</sup> *Department of Cardiology, Nancy University Hospital, 54511 Vandoeuvre-lès-Nancy, France*

<sup>v</sup> *Clinique Pasteur, 31076 Toulouse, France*

Corresponding author at: Institut de Cardiologie, Groupe Hospitalier Pitié-Salpêtrière, 47–83 boulevard de l'Hôpital, 75013 Paris, France.

*E-mail address: [estelle.gandjbakhch@aphp.fr](mailto:estelle.gandjbakhch@aphp.fr) (E. Gandjbakhch).*

Estelle Gandjbakhch, Jean Nicolas Dacher, Alexis Jacquier and Serge Boveda contributed equally to this work.

## **Summary**

Magnetic resonance imaging (MRI) has become the reference imaging technique for the management of a large number of diseases. The number of MRI examinations increases every year, simultaneously with the number of patients receiving a cardiac electronic implantable device (CEID). The presence of a CEID was considered an absolute contraindication for MRI for many years. The progressive replacement of conventional pacemakers and defibrillators by “magnetic resonance (MR)-conditional” CEIDs and recent data on the safety of MRI in patients with “MR-non-conditional” CEIDs have gradually increased the demand for MRI in patients with a CEID. However, some risks are associated with MRI in CEID carriers, even with MR-conditional devices, because these devices are not “MR safe”. Specific programming of the device in “MR mode” and monitoring patients during MRI remain mandatory for all patients with a CEID. A standardized patient workflow based on an institutional protocol should be established in each institution performing such examinations. This joint position paper of the Working Group of Pacing and Electrophysiology of the French Society of Cardiology and the French Society of Diagnostic and Interventional Cardiac and Vascular Imaging describes the effect of and risks associated with MRI in CEID carriers. We propose recommendations for patient workflow and monitoring and CEID programming in MR-conditional, “MR-conditional non-guaranteed” and MR-non-conditional devices.

## **Résumé**

L'imagerie par résonance magnétique (IRM) est devenue l'imagerie de référence pour le diagnostic d'un grand nombre de pathologies. Le nombre d'examens IRM est en augmentation constante, parallèlement au nombre de patients implantés avec un dispositif électronique cardiaque implantable (DECI). L'IRM était considérée comme une contre-indication absolue à l'IRM jusqu'à il y a quelques années. Le remplacement progressif des stimulateurs et des défibrillateurs conventionnels par des dispositifs « IRM compatibles sous conditions » ainsi que les données récentes sur la sécurité des examens IRM chez les patients porteurs de dispositifs « non-IRM compatible » ont considérablement accru la demande d'examens IRM chez les patients porteurs de DECI. Cependant, les IRM peuvent être associés à certains risques chez les porteurs de DECI, y compris ceux « IRM compatibles sous conditions ». La programmation spécifique de l'appareil en mode « IRM » et la surveillance des patients pendant l'examen restent obligatoires pour tous les patients porteurs de CEDI. Une prise en

charge standardisée s'appuyant sur un protocole institutionnel est recommandée dans chaque établissement effectuant de tels examens. Ce document conjoint du Groupe Rythmologie et Stimulation Cardiaque de la Société Française de Cardiologie et de la Société Française d'Imagerie Cardiaque et Vasculaire Diagnostique et Interventionnelle décrit les effets et les risques associés à l'IRM chez des porteurs de DECI. Nous proposons des recommandations pour la prise en charge de ces patients, leur surveillance, ainsi que la programmation des dispositifs qu'ils soient « IRM compatibles sous conditions », « IRM compatibles sous conditions hors garantie constructeur » ou « non-IRM compatibles ».

## **KEYWORDS**

Magnetic resonance imaging;  
Cardiac electronic implantable device;  
Pacemaker;  
Defibrillator;  
French Society of Cardiology

## **MOTS CLÉS**

Imagerie par résonance magnétique ;  
Dispositif électronique cardiaque implantable ;  
Stimulateur cardiaque ;  
Défibrillateur ;  
Société Française de Cardiologie ;

*Abbreviations:* CEID, cardiac electronic implantable device; ICD, implantable cardioverter defibrillator; MR, magnetic resonance; MRI, magnetic resonance imaging.

## Background

The rate of cardiac electronic implantable device (CEID) implantation is increasing every year. An estimated 4 million patients carry a CEID worldwide. Each year, more than 500,000 pacemakers and 85,000 implantable cardioverter defibrillators (ICDs) are implanted in Europeans (European Heart Rhythm Association data [1]). In France, about 400,000 patients carry a CEID, with approximately 70,000 pacemakers and 15,000 ICDs implanted in 2018 (International Health Market Trends data). At least one in 50 people aged  $\geq 75$  years will have a permanent pacemaker implanted [2]. At the same time, magnetic resonance imaging (MRI) has become the reference imaging technique for the management of a large number of diseases, and the number of examinations performed increases every year (+12%), with approximately 7 million MRI examinations in France in 2017 [3].

For many years, the presence of a CEID, such as a pacemaker or an ICD, was considered an absolute contraindication for MRI. Two major developments have changed this paradigm in recent years. First, manufacturers have progressively marketed new “MR-conditional” systems. However, these MR-conditional materials are not “MR safe”, and therefore require specific device programming and patient monitoring. Second, several large observational studies have shown that MRI can also be performed in patients carrying an “MR-non-conditional” CEID with a low risk of complications, which shifts the presence of an MRI non-conditional CEID from an absolute to a relative contraindication. As a class IIb, level B recommendation, the 2013 European Society of Cardiology guidelines on pacing and resynchronization therapy allow for MRI with a conventional MR-non-conditional CEID if appropriate precautions are taken [4]. In 2017, the Heart Rhythm Society expert consensus statement on MRI and radiation exposure in patients with CIEDs issued a class IIa, level B recommendation for this indication [5]. However, for all patients carrying an MR-conditional or MR-non-conditional device, any MRI should be integrated into a standardized workflow, defined in an institutional protocol involving both MRI specialists and device specialists [6].

Despite these recommendations, MRI remains underused in patients carrying a CEID. A patient with an ICD is 50 times less likely to benefit from MRI than a patient without implantation [7]. The reasons are multiple: issues related to local organization; the difficulty of establishing a concerted institutional workflow; the availability of device specialists; legal/responsibility issues between MRI specialists and cardiologists; the unjustified fear of some patients or treating physicians because of

lack of knowledge of the recent recommendations; and the lack of financial recognition of the complexity of MRI in CEID carriers.

This position paper gives the common position of the Working Group of Pacing and Electrophysiology of the French Society of Cardiology (Société Française de Cardiologie; SFC) and the French Society of Diagnostic and Interventional Cardiac and Vascular Imaging (Société Française d'Imagerie Cardiaque et Vasculaire Diagnostique et Interventionnelle; SFICV) on the technical conditions for MRI in patients with MR-conditional and MR-non-conditional CEIDs, which could serve as a basis for institutional MRI protocols in patients with a CEID. This consensus was based on an extensive analysis of the current literature, followed by exchanges between CEID and MRI specialists representing both societies.

## **Definitions**

An MR-conditional CEID is defined as a whole system – consisting of a generator, MR protection mode software and leads – that has been tested and approved by manufacturers for MRI under specific conditions of use. Modifications have been made to the material to limit the effect of the magnetic and radiofrequency fields on the device and the patient. Only systems combining leads and generators from the same manufacturer have been specifically tested to be safe and are guaranteed by the manufacturer as MR conditional. Specific MR-mode programming is always required during the MRI to limit the effect of the magnetic and radiofrequency fields on the functioning of the device. All MR-conditional systems exclude epicardial devices, abandoned or fractured leads and lead extensions/adapters.

MR-conditional non-guaranteed CEIDs are defined as systems consisting of MR-conditional generators and leads issued by different manufacturers. MR-non-conditional CEIDs are all other devices.

An updated list of MR-conditional CEIDs is provided at [www.irm-compatibilite.com](http://www.irm-compatibilite.com); the list was created with the support of the Working Group of Pacing and Electrophysiology of the French Society of Cardiology.

Pacing-dependent patients are defined as those with an inadequate or even absent intrinsic rhythm (i.e. asystole longer than 5 s or spontaneous frequency of < 30/min) [8]. Patients with

permanent bradycardia are defined as those with permanent spontaneous cardiac frequency < 50/min.

“On site” means within the same hospital. “On the premises” means within the same building.

## **Effect of MRI on CEIDs**

During MRI, three magnetic fields are involved: a static field called B0 (1.5–3 T in current magnet technology, but higher fields are now commercially available); a three-dimensional gradient magnetic field (Gx, Gy, Gz); and a radiofrequency (B1) field. All three fields can interfere with the functioning of the device. The different risks associated with MRI are shown in [Table 1](#). The static magnetic field B0 can theoretically induce force and torque to the ferromagnetic components that are present within the generator (none is present in conventional leads), but movement of the generator is unlikely [9]. The mechanical switch of MR-non-conditional generators can be activated by MRI, thereby resulting in asynchronous pacing in pacemakers or deactivation of tachycardia detection in ICDs [10]. All magnetic fields can cause electrical reset of MR-non-conditional generators, leading to backup in emergency mode with VVI pacing and reactivation of therapies that could cause pacing inhibition or inappropriate shocks [11-13].

Rapid depletion of the battery can also occur [14]. A gradient magnetic field can induce a current within the conductive wire of the lead that can lead to induced myocardial capture. The gradient and B1 (radiofrequency) magnetic fields can generate oversensing that can lead to pacing inhibition or inappropriate therapies [13]. MR-non-conditional leads can receive the B1 (radiofrequency) field as an antenna and transmit the energy to the myocardium, thereby generating arrhythmias, tissue heating and damage around the lead, leading to increased capture threshold or decreased sensing [15-17]. This risk appears to be particularly great with abandoned leads.

Some risks are associated with the temporary MR mode, and are common to MR-conditional and MR-non-conditional devices. During MRI, an asynchronous mode (VOO/DOO) or a deactivation of pacing mode (ODO/OOO), according to the underlying rhythm of the patient, and deactivation of therapy detection in ICDs should be programmed to avoid oversensing, leading to pacing inhibition or inappropriate therapies. In asynchronous mode, there is a very low risk (< 1/10,000) of induced ventricular arrhythmia caused by inappropriate pacing in a ventricular vulnerable period [18]. This complication has been mainly described in patients with low left ventricular ejection fraction, acute



coronary syndrome or hydroelectrolyte disturbances and in patients who are not pacing dependent [19]. For ICDs, the deactivation of tachycardia detection carries the risk that a ventricular arrhythmia cannot be treated during this time. In patients who are not pacing dependent and are programmed in ODO/ODO, there is a risk of acute bradycardia. Although all the mentioned risks seem very low, they remain difficult to assess, and are unpredictable at the patient level.

There have been some concerns about the risk of thoracic and cardiac MRI in patients with CEIDs because of the close proximity of the device. However, most studies have shown a similar safety profile between cardiac/thoracic and extrathoracic MRI [20-22]. CEIDs, especially ICDs [20] and CEIDs positioned at the left side [21, 23], can cause artefacts. Cardiac artefacts caused by the device can be a concern, but specific techniques (frequency-scout acquisitions, spoiled gradient echo, reduced echo time, fast spin echo) may reduce the artefacts [24].

### **General conditions for MRI in patients with a CEID**

As stated above, CEIDs are not MR safe, but are rather MR-conditional materials. Thus, a standardized patient workflow needs to be established by each institution, based on an institutional protocol decided with consensus between MRI specialists and device specialists. This workflow should include the benefit/risk ratio of the MRI (particularly in patients with MR-non-conditional devices), evaluation of a possible alternative imaging modality (frequently computed tomography) and the exclusion of patient- or device-related contraindications ([Table 2](#)). The risk from MRI in a patient with an implant is considerably lower than that from device removal before MRI [25]. For all patients, one should check the precise characteristics of the material (manufacturer and models of generators and all leads), the medical indication of the device, the underlying rhythm of the patient and whether they are pacing dependent, as well as the history of ventricular arrhythmias in ICD carriers. A transmission form including all information needed before the MRI examination is proposed in [Fig. 1](#) and [Fig. A.1](#).

Despite some evidence that MRI within the first weeks of implantation is safe [26, 27], we recommend in the absence of an emergency to respect a 6-week delay after CEID implantation. Epicardial, fractured or abandoned leads as well as adapters and lead extensions are classical contraindications for MRI and, in some cases, a chest radiograph can be performed to exclude them. For all devices, one should verify the integrity of the device (generator and leads) before the MRI (lead

impedance, capture voltage threshold, sensing and battery status); MRI is contraindicated in case of an elective replacement indicator.

During the MRI examination, we recommend that all patients with MR-conditional or MR-non-conditional devices be at least monitored with cardiac frequency from pulse oximetry. If possible, electrocardiogram monitoring and visual/voice contact with a physician or qualified staff member are advised. Because MR sequences can cause electrocardiogram artefacts, monitoring of cardiac frequency with pulse oximetry is mandatory for all patients. Although electrocardiogram monitoring is advised in addition to pulse oximetry, it is not mandatory if cardiac frequency can be efficiently monitored with pulse oximetry. An external defibrillator and emergency materials should be present on site. Physicians with the ability to perform resuscitation and advanced cardiac life support should be available immediately on an emergency standby basis, as defined by the institutional protocol. Physicians with device programming skills should be available on an emergency standby basis, depending on the conditions defined by the institutional protocol (Fig. 2 and Fig. 3).

### **Workflow for MR-conditional CEIDs**

MR-conditional CEIDs have been tested and approved (CE certification) for MRI under specific conditions. MR-conditional generators and leads have been modified by manufacturers to limit the influence of magnetic and radiofrequency fields on the system. The safety of MRI has been validated in clinical trials for some systems: EnRhythm SureScan® and Advisa® pacemakers and Evera® ICD (Medtronic Inc., Minneapolis, MN, USA); Entovis ProMRI® and Evia® pacemakers and Iforia® ICD (Biotronik Inc., Lake Oswego, OR, USA); and Kora® pacemaker (Microport, Shanghai, China) [28-35]. Because clinical validation is limited by practical/logistical issues and does not allow for validating thousands of variables that could affect ICD or pacemaker systems during MRI, MR-conditional materials are now validated by computer modelling that enables a large number of conditions to be tested [36]. On the basis of these tests, each manufacturer provides specific guidelines and conditions under which the safety of the MRI is guaranteed. Some systems have been validated for only 1.5 T, others for 3 T; some include a thoracic exclusion zone whereas others allow full-body MRI. Hence, these specific conditions and guidelines can vary between manufacturers, and can only be applied to a whole validated system (i.e. generator plus leads). The specific manufacturer recommendations for each system are available at each manufacturer's website or at [www.irm-compatibilite.com](http://www.irm-compatibilite.com).

Before the MRI, the system should be validated as MR conditional by the physician. The workflow for MR-conditional devices should be assessed in a standardized institutional protocol following the general recommendations specified above. The time and location of the reprogramming of the device before an MRI depend mainly on the potential impact of the temporary MR mode on patient safety: lack of pacing of acute bradycardia in ODO/ODO mode; triggered ventricular arrhythmia with asynchronous pacing; or lack of treatment of ventricular arrhythmias with ICDs. This risk increases with time when the temporary programme is active. In patients who are not pacing dependent, programming the device in an inhibited pacing mode (VVI/DDI) seems safe, although this programming is off manufacturer guarantee. Inhibited mode in patients who are not pacing dependent decreases the risk of a non-treated paroxysmal bradycardia and the risk associated with asynchronous pacing [22, 37]. Inhibited modes may be preferred with paroxysmal atrioventricular block/sinus node dysfunction (off manufacturer guarantee) [22].

The reprogramming of devices before and after MRI could reasonably be performed on the same day as the MRI, on site, but at a different place from the MRI scan (cardiology outpatient clinic) (Fig. 2). We recommend that the time for which the patient remains on MR mode should be as short as possible, to limit the risks associated with lack of pacing/therapy or asynchronous pacing.

The reprogramming of devices before and after MRI could reasonably be performed just before and after the MRI in high-risk patients with an unstable clinical cardiac condition or with recent (< 15 days) ICD therapy (Fig. 2).

Because several studies have shown that MRI is safe for MR-conditional devices, the presence of the device specialist during the MRI scan is not mandatory. However, a device specialist should be available on call, as specified in the institutional protocol.

Some devices have a specific algorithm allowing for the automatic detection of an MR field, leading to the automatic activation of the temporary prespecified MR programme. In these cases, the temporary MR mode will be activated only during the MRI, and baseline settings will be restored automatically at the end of the examination. For these devices, the control and programming of the device by the device specialist can be performed several days before the scan.

The maximum MR field and exclusion zone conditions should be applied according to the manufacturer's recommendations (available at each manufacturer's website or at <http://irm-compatibilite.com>).

## **Workflow for MR-conditional non-guaranteed CEIDs**

A current issue is to determine in which category to include the MR-conditional non-guaranteed CEIDs, which are defined by MR-conditional materials (leads and generators), but from different manufacturers. By definition, MR-conditional generators have been validated only in combination with the MR-conditional leads from the same manufacturer. No published data have specifically addressed this issue. However, from expert experience, the MR-conditional CEID workflow could reasonably be applied to these devices (Fig. 2). A national registry should be developed for these patients to validate the safety of this workflow applied to MR-conditional non-guaranteed CEIDs.

## **Workflow for MR-non-conditional CEIDs**

Recent clinical observational retrospective and prospective data have demonstrated the relative safety of MRI in patients with MR-non-conditional material. In the MagnaSafe registry, including extrathoracic MRI performed with 1000 MR-non-conditional pacemakers and 500 ICDs, six cases of electrical reset, one case of generator heating and six cases of atrial arrhythmias were observed [27]. No ventricular arrhythmia occurred. One patient without adequate MR programming presented ICD generator dysfunction requiring immediate replacement. Minor increases in voltage capture threshold or lead impedances, decreased sensing or battery depletions have been observed in 0.4–4% of cases, but none led to the loss of capture, programming changes or generator/lead replacement. Pacing-dependent patients with ICDs, epicardial or abandoned leads or generators with an elective replacement indicator were excluded. Although the risk of complications appears low, it seems unpredictable, and could have substantial consequences for pacing-dependent patients.

Allowing for extrathoracic MRI in patients with MR-non-conditional CEIDs seems reasonable if MRI is the more accurate test for the patient's condition. MRI indications should be evaluated on a risk/benefit balance basis for each patient, especially pacing-dependent patients. Information on the risk associated with the MRI should be provided to the patient. The workflow for MR-non-conditional devices should be assessed in a standardized institutional protocol following the general recommendations specified above. We recommend monitoring (cardiac frequency by pulse oximetry ± electrocardiogram monitoring, visual contact) of all patients with MR-non-conditional devices during MRI in the presence of a physician or qualified and trained staff member. For pacing-dependent

patients, physicians with skills in programming devices should be present or available immediately on an emergency standby basis during the MRI examination. For patients who are not pacing dependent, a device specialist should be available immediately on the premises, as defined by the institutional protocol. The reprogramming of devices before and after MRI could reasonably be performed on the same day as the MRI, on site, but at a different place from the MRI scan (cardiology outpatient clinic) (Fig. 3). We recommend that the time for which the patient remains on MR mode should be as short as possible, to limit the risks associated with absence of pacing/therapy or asynchronous pacing.

The reprogramming of devices before and after MRI could reasonably be performed on the premises just before and after the MRI examination in high-risk patients with an unstable clinical cardiac condition or with recent (< 15 days) ICD therapy (Fig. 3).

Because of very low evidence of MR safety for MRI scanning > 1.5 T, we recommend limiting the MR field strength to 1.5 T for non-conditional devices [38]. To limit the risk of conducting radiofrequency pulses to the myocardium within the conductive lead, we recommend limiting the whole-body specific absorption rate to a minimum. A specific absorption rate < 3.2 W/kg for head examinations and 2 W/kg for body examinations are commonly advised. We advise scanning in standard mode and avoiding specific absorption rate levels 1 and 2. We also recommend limiting the time of exposure and number of sequences to those absolutely necessary.

Control of the device and restoration of the baseline settings should be performed as soon as possible after the end of the MRI scan. If significant modification of lead parameters is observed (increase of capture threshold voltage > 0.5V/0.4 ms, decrease of sensing > 50%, modification of impedance > 100 ohms or high-voltage lead impedance > 10 ohms), remote monitoring or early follow-up within 2 weeks after the MRI scan is recommended.

Cardiac MRI may be associated with increased risk of interference because of the location of device inside the radiofrequency field. However, with MR-non-conditional CEIDs, the indication for cardiac MRI should be discussed between the referring cardiologist, the device specialist and the MRI specialist. Cardiac MRI should be restricted to indications for which alternative methods are inaccurate and only performed in experienced centres. Cardiac CT may be used as an alternative, when suitable.

## **Epicardial and abandoned leads**

We have little data on MR safety in patients carrying epicardial, fractured or abandoned leads, because these patients were excluded from observational studies. In small case series, no complication of MRI was observed in these patients [39-43]. However, we think that these data are insufficient to recommend MRI in these cases, and that the presence of epicardial, fractured or abandoned leads should remain a contraindication for MRI. In individual cases with a life-threatening emergency, non-thoracic MRI can be discussed in patients who are not pacing dependent after careful consideration of the benefit/risk ratio and multidisciplinary discussion.

### **Implantable loop recorders**

Implantable loop recorders are MR safe. No specific MR-mode programming is necessary before MRI, and no monitoring of the patient is advised. MRI can cause artefacts that can be recorded by the device and overload the memory. The patient should notify their treating device specialist of any MRI that occurred during follow-up. However, to avoid any problems in radiology departments, any MRI requested for a patient with an implantable loop recorder should mention the device and its full compatibility with MRI.

### **Subcutaneous ICDs and leadless pacemakers**

The first-generation subcutaneous ICD (SQ-RX™; Boston Scientific, Marlborough, MA, USA) was not labelled as MR conditional, but the second- and third-generation subcutaneous ICDs (EMBLEM™ models A209 and A219; Boston Scientific) are guaranteed to be MR conditional (3 T, full body) [44]. The leadless pacemaker available on the market (Micra™; Medtronic) is also guaranteed to be MR conditional (3 T, full body) [45]. The same workflow as for conventional MR-conditional material should be applied to these devices. However, cardiac imaging can be affected by subcutaneous ICDs and leadless pacemakers, mostly because metallic artefacts on the left ventricle can prevent accurate tissue characterization [44].

### **Sources of funding**

None.

### **Disclosure of interest**

The authors declare that they have no conflicts of interest concerning this article.

## References

- [1] Raatikainen MJ, Arnar DO, Zeppenfeld K, et al. Statistics on the use of cardiac electronic devices and electrophysiological procedures in the European Society of Cardiology countries: 2014 report from the European Heart Rhythm Association. *Europace* 2015;17 Suppl 1:i1-75.
- [2] Bradshaw PJ, Stobie P, Knuiman MW, Briffa TG, Hobbs MS. Trends in the incidence and prevalence of cardiac pacemaker insertions in an ageing population. *Open Heart* 2014;1:e000177.
- [3] OECD. Magnetic resonance imaging (MRI) exams (indicator). 2019. Available at: <https://data.oecd.org/healthcare/magnetic-resonance-imaging-mri-exams.htm>.
- [4] Brignole M, Auricchio A, Baron-Esquivias G, et al. 2013 ESC guidelines on cardiac pacing and cardiac resynchronization therapy: the task force on cardiac pacing and resynchronization therapy of the European Society of Cardiology (ESC). Developed in collaboration with the European Heart Rhythm Association (EHRA). *Europace* 2013;15:1070-118.
- [5] Indik JH, Gimbel JR, Abe H, et al. 2017 HRS expert consensus statement on magnetic resonance imaging and radiation exposure in patients with cardiovascular implantable electronic devices. *Heart Rhythm* 2017;14:e97-e153.
- [6] Cruyppeninck Y, Dubourg B, Michelin P, et al. Pacemakers and MRI: A protocol in line with international guidelines and approved by the SFICV (French Society of Cardiovascular Imaging). *Diagn Interv Imaging* 2017;98:203-15.
- [7] Nazarian S, Reynolds MR, Ryan MP, Wolff SD, Mollenkopf SA, Turakhia MP. Utilization and likelihood of radiologic diagnostic imaging in patients with implantable cardiac defibrillators. *J Magn Reson Imaging* 2016;43:115-27.
- [8] Korantzopoulos P, Letsas KP, Grekas G, Goudevenos JA. Pacemaker dependency after implantation of electrophysiological devices. *Europace* 2009;11:1151-5.
- [9] Luechinger R, Duru F, Scheidegger MB, Boesiger P, Candinas R. Force and torque effects of a 1.5-Tesla MRI scanner on cardiac pacemakers and ICDs. *Pacing Clin Electrophysiol* 2001;24:199-205.
- [10] Luechinger R, Duru F, Zeijlemaker VA, Scheidegger MB, Boesiger P, Candinas R. Pacemaker reed switch behavior in 0.5, 1.5, and 3.0 Tesla magnetic resonance imaging units: are reed



- switches always closed in strong magnetic fields? *Pacing Clin Electrophysiol* 2002;25:1419-23.
- [11] Gimbel JR. Unexpected asystole during 3T magnetic resonance imaging of a pacemaker-dependent patient with a 'modern' pacemaker. *Europace* 2009;11:1241-2.
- [12] Higgins JV, Sheldon SH, Watson RE, Jr., et al. "Power-on resets" in cardiac implantable electronic devices during magnetic resonance imaging. *Heart Rhythm* 2015;12:540-4.
- [13] Maffe S, Paffoni P, Perucca A, Signorotti F, Dellavesa P, Parravicini U. Pseudo "end of life" indication after electromagnetic field exposure: an unusual effect of magnetic resonance imaging on implanted cardioverter defibrillator. *Int J Cardiol* 2012;156:e36-9.
- [14] Zaremba T, Jakobsen AR, Thogersen AM, Oddershede L, Riahi S. The effect of radiotherapy beam energy on modern cardiac devices: an in vitro study. *Europace* 2014;16:612-6.
- [15] Fontaine JM, Mohamed FB, Gottlieb C, Callans DJ, Marchlinski FE. Rapid ventricular pacing in a pacemaker patient undergoing magnetic resonance imaging. *Pacing Clin Electrophysiol* 1998;21:1336-9.
- [16] Horwood L, Attili A, Luba F, et al. Magnetic resonance imaging in patients with cardiac implanted electronic devices: focus on contraindications to magnetic resonance imaging protocols. *Europace* 2017;19:812-7.
- [17] Luechinger R, Zeijlemaker VA, Pedersen EM, et al. In vivo heating of pacemaker leads during magnetic resonance imaging. *Eur Heart J* 2005;26:376-83; discussion 25-7.
- [18] Nowak B. [Is asynchronous ventricular pacemaker stimulation dangerous? Results of an international survey]. *Dtsch Med Wochenschr* 2005;130:997-1001.
- [19] Irnich W, Irnich B, Bartsch C, Stertmann WA, Gufler H, Weiler G. Do we need pacemakers resistant to magnetic resonance imaging? *Europace* 2005;7:353-65.
- [20] Buendia F, Cano O, Sanchez-Gomez JM, et al. Cardiac magnetic resonance imaging at 1.5 T in patients with cardiac rhythm devices. *Europace* 2011;13:533-8.
- [21] Naehle CP, Kreuz J, Strach K, et al. Safety, feasibility, and diagnostic value of cardiac magnetic resonance imaging in patients with cardiac pacemakers and implantable cardioverters/defibrillators at 1.5 T. *Am Heart J* 2011;161:1096-105.
- [22] Nazarian S, Hansford R, Rahsepar AA, et al. Safety of Magnetic Resonance Imaging in Patients with Cardiac Devices. *N Engl J Med* 2017;377:2555-64.

- [23] Sasaki T, Hansford R, Zviman MM, et al. Quantitative assessment of artifacts on cardiac magnetic resonance imaging of patients with pacemakers and implantable cardioverter-defibrillators. *Circ Cardiovasc Imaging* 2011;4:662-70.
- [24] Hilbert S, Jahnke C, Loebe S, et al. Cardiovascular magnetic resonance imaging in patients with cardiac implantable electronic devices: a device-dependent imaging strategy for improved image quality. *Eur Heart J Cardiovasc Imaging* 2018;19:1051-61.
- [25] Bongiorno MG, Kennergren C, Butter C, et al. The European Lead Extraction ConTRolled (ELECTRa) study: a European Heart Rhythm Association (EHRA) Registry of Transvenous Lead Extraction Outcomes. *Eur Heart J* 2017;38:2995-3005.
- [26] Friedman HL, Acker N, Dalzell C, et al. Magnetic resonance imaging in patients with recently implanted pacemakers. *Pacing Clin Electrophysiol* 2013;36:1090-5.
- [27] Russo RJ, Costa HS, Silva PD, et al. Assessing the Risks Associated with MRI in Patients with a Pacemaker or Defibrillator. *N Engl J Med* 2017;376:755-64.
- [28] Awad K, Griffin J, Crawford TC, et al. Clinical safety of the Iforia implantable cardioverter-defibrillator system in patients subjected to thoracic spine and cardiac 1.5-T magnetic resonance imaging scanning conditions. *Heart Rhythm* 2015;12:2155-61.
- [29] Bailey WM, Mazur A, McCotter C, et al. Clinical safety of the ProMRI pacemaker system in patients subjected to thoracic spine and cardiac 1.5-T magnetic resonance imaging scanning conditions. *Heart Rhythm* 2016;13:464-71.
- [30] Bailey WM, Rosenthal L, Fananapazir L, et al. Clinical safety of the ProMRI pacemaker system in patients subjected to head and lower lumbar 1.5-T magnetic resonance imaging scanning conditions. *Heart Rhythm* 2015;12:1183-91.
- [31] Gimbel JR, Bello D, Schmitt M, et al. Randomized trial of pacemaker and lead system for safe scanning at 1.5 Tesla. *Heart Rhythm* 2013;10:685-91.
- [32] Gold MR, Sommer T, Schwitter J, et al. Full-Body MRI in Patients With an Implantable Cardioverter-Defibrillator: Primary Results of a Randomized Study. *J Am Coll Cardiol* 2015;65:2581-8.
- [33] Savoure A, Mechulan A, Burban M, Olivier A, Lazarus A. The Kora Pacemaker is Safe and Effective for Magnetic Resonance Imaging. *Clin Med Insights Cardiol* 2015;9:85-90.

- [34] Shenthar J, Milasinovic G, Al Fagih A, et al. MRI scanning in patients with new and existing CapSureFix Novus 5076 pacemaker leads: randomized trial results. *Heart Rhythm* 2015;12:759-65.
- [35] Wilkoff BL, Bello D, Taborsky M, et al. Magnetic resonance imaging in patients with a pacemaker system designed for the magnetic resonance environment. *Heart Rhythm* 2011;8:65-73.
- [36] Wilkoff BL, Albert T, Lazebnik M, et al. Safe magnetic resonance imaging scanning of patients with cardiac rhythm devices: a role for computer modeling. *Heart Rhythm* 2013;10:1815-21.
- [37] Sommer T, Bauer W, Fischbach K, et al. MR Imaging in Patients with Cardiac Pacemakers and Implantable Cardioverter Defibrillators. *Rofo* 2017;189:204-17.
- [38] Naehle CP, Meyer C, Thomas D, et al. Safety of brain 3-T MR imaging with transmit-receive head coil in patients with cardiac pacemakers: pilot prospective study with 51 examinations. *Radiology* 2008;249:991-1001.
- [39] Higgins JV, Gard JJ, Sheldon SH, et al. Safety and outcomes of magnetic resonance imaging in patients with abandoned pacemaker and defibrillator leads. *Pacing Clin Electrophysiol* 2014;37:1284-90.
- [40] Kanal E. Safety of MR imaging in patients with retained epicardial pacer wires. *AJR Am J Roentgenol* 1998;170:213-4.
- [41] Langman DA, Goldberg IB, Finn JP, Ennis DB. Pacemaker lead tip heating in abandoned and pacemaker-attached leads at 1.5 Tesla MRI. *J Magn Reson Imaging* 2011;33:426-31.
- [42] Pulver AF, Puchalski MD, Bradley DJ, et al. Safety and imaging quality of MRI in pediatric and adult congenital heart disease patients with pacemakers. *Pacing Clin Electrophysiol* 2009;32:450-6.
- [43] Strach K, Naehle CP, Muhlsteffen A, et al. Low-field magnetic resonance imaging: increased safety for pacemaker patients? *Europace* 2010;12:952-60.
- [44] Keller J, Neuzil P, Vymazal J, et al. Magnetic resonance imaging in patients with a subcutaneous implantable cardioverter-defibrillator. *Europace* 2015;17:761-6.
- [45] Blessberger H, Kiblboeck D, Reiter C, et al. Monocenter Investigation Micra(R) MRI study (MIMICRY): feasibility study of the magnetic resonance imaging compatibility of a leadless pacemaker system. *Europace* 2019;21:137-41.

## Figure legends

**Figure 1.** Transmission form. ICD: implantable cardioverter defibrillator; ID: identification; LV: left ventricular; MR: magnetic resonance; MRI: magnetic resonance imaging; NA: not available; ODO/OOO: atrium and ventricle being sensed, no pacing (ODO)/deactivation of sensing and pacing (OOO); RV: right ventricular; SFICV: Société Française d'Imagerie Cardiaque et Vasculaire Diagnostique et Interventionnelle (French Society of Diagnostic and Interventional Cardiac and Vascular Imaging); VOO/DOO: ventricular pacing (VOO) or dual chamber pacing (atrium and ventricle) (DOO), no sensing (asynchronous mode); VVI/DDI: ventricular pacing and ventricular sensing, inhibition of a sensed beat (VVI)/dual chamber pacing and dual chamber sensing, inhibition of a sensed beat (DDI).

**Figure 2.** Workflow for magnetic resonance (MR)-conditional guaranteed and non-guaranteed cardiac electronic implantable devices (CEIDs). AVB: atrioventricular block; ICD: implantable cardioverter defibrillator; MRI: magnetic resonance imaging; ODO/OOO: atrium and ventricle being sensed, no pacing (ODO)/deactivation of sensing and pacing (OOO); SND: sinus node dysfunction; VOO/DOO: ventricular pacing (VOO) or dual chamber pacing (atrium and ventricle) (DOO), no sensing (asynchronous mode); VVI/DDI: ventricular pacing and ventricular sensing, inhibition of a sensed beat (VVI)/dual chamber pacing and dual chamber sensing, inhibition of a sensed beat (DDI).

<sup>a</sup> Except for devices with automatic detection of MR field.

**Figure 3.** Workflow for magnetic resonance (MR)-non-conditional cardiac electronic implantable devices (CEIDs). AVB: atrioventricular block; ICD: implantable cardioverter defibrillator; MRI: magnetic resonance imaging; ODO/OOO: atrium and ventricle being sensed, no pacing (ODO)/deactivation of sensing and pacing (OOO); SND: sinus node dysfunction; VOO/DOO: ventricular pacing (VOO) or dual chamber pacing (atrium and ventricle) (DOO), no sensing (asynchronous mode); VVI/DDI: ventricular pacing and ventricular sensing, inhibition of a sensed beat (VVI)/dual chamber pacing and dual chamber sensing, inhibition of a sensed beat (DDI).

<sup>a</sup> In pacing-dependent patient, control device as soon as possible.

**Figure A.1.** Transmission form in French. DAI: implantable cardioverter defibrillator; VG: left ventricular; IRM: magnetic resonance imaging; ND: not available; ODO/OOO: atrium and ventricle being sensed, no pacing (ODO)/deactivation of sensing and pacing (OOO); VD: right ventricular; SFICV: Société Française d'Imagerie Cardiaque et Vasculaire Diagnostique et Interventionnelle (French Society of Diagnostic and Interventional Cardiac and Vascular Imaging); VOO/DOO: ventricular pacing (VOO) or dual chamber pacing (atrium and ventricle) (DOO), no sensing (asynchronous mode); VVI/DDI: ventricular pacing and ventricular sensing, inhibition of a sensed beat (VVI)/dual chamber pacing and dual chamber sensing, inhibition of a sensed beat (DDI).

**Table 1** Risks associated with magnetic resonance imaging in patients with magnetic resonance-non-conditional and magnetic resonance-conditional cardiac electronic implantable devices.

MR-non-conditional CEIDs	<p>Acute bradycardia in ODO/OOO mode</p> <p>Inactivation of ICD therapy: absence of VT/VF treatment</p> <p>Oversensing → pacing inhibition/inappropriate ICD therapy</p> <p>Ventricular arrhythmia induced by asynchronous pacing mode (VOO/DOO)</p> <p>Power on reset mode and emergency mode (usually VVI with risk of pacing inhibition by pulsed MR fields and risk of reactivation of ICD therapies)</p> <p>Reed switch → asynchronous pacing/inhibition of tachycardia detection</p> <p>Transmission of radiofrequency field: tissue heating and damage, arrhythmias, change in capture or sensing thresholds</p> <p>Battery depletion</p> <p>Gradient magnetic field-induced electrical current → oversensing, myocardial rapid capture, arrhythmias</p> <p>Magnetic-induced force and torque (generator)</p>
MR-conditional CEIDs under specific conditions	<p>Acute bradycardia in ODO/OOO mode</p> <p>Inactivation of ICD therapy: absence of VT/VF treatment</p> <p>Oversensing → pacing inhibition/inappropriate ICD therapy</p> <p>Ventricular arrhythmia induced by asynchronous pacing mode (VOO/DOO)</p>

CEID: cardiac electronic implantable device; ICD: implantable cardiac defibrillator; MR: magnetic resonance; ODO/OOO: atrium and ventricle being sensed

---

(ODO)/deactivation of CEID (OOO); VF: ventricular fibrillation; VOO/DOO: ventricular pacing (VOO) or dual chamber pacing (atrium and ventricle) (DOO) (asynchronous mode); VT: ventricular tachycardia; VVI: ventricular pacing and ventricular sensing, inhibition of a sensed beat.

**Table 2** Common workflow for magnetic resonance imaging in patients with magnetic resonance-conditional and magnetic resonance-non-conditional cardiac electronic implantable devices.

Before MRI scan	<p>Validate the clinical benefit of the MRI scan (consider possible alternative imaging)</p> <p>Verify integrity of the system (battery, leads)</p> <p>Characteristics of the device (date of implantation, manufacturer and model of generator and leads): MR-conditional or MR-non-conditional system?</p> <p>Medical indication of the device, pacing dependency, history of ventricular arrhythmias</p> <p>Exclude contraindications: epicardial, fractured and abandoned leads as well as adapters and lead extensions (with X-ray if necessary); high capture thresholds &gt; 2 V/0.4 ms; out-of-range impedance values &lt; 200 or &gt; 1500 ohms; elective replacement indicator or end of service</p> <p>Set specific MR pacing programme according to the underlying rhythm, deactivate tachycardia detection (ICDs)</p>
During MRI scan	<p>Monitoring (cardiac frequency by pulse oximetry + electrocardiogram monitoring if possible + visual monitoring) by physician or qualified personnel</p> <p>Presence of a defibrillator and emergency material</p> <p>Physicians with the skill to perform resuscitation available immediately</p> <p>Physicians with the skill to programme devices available on call or immediately, depending on the device and patient dependency (<a href="#">Fig. 2</a> and <a href="#">Fig. 3</a>)</p>
After MRI scan	<p>Device control (battery, sensing, impedance, pacing threshold) and reprogramming of baseline settings, reactivation of tachycardia</p>



detection (ICDs)

---

ICD: implantable cardiac defibrillator; MR: magnetic resonance; MRI: magnetic resonance imaging.

Patient name: .....

ID
----

Indication for device implantation: .....

**GENERATOR**

Date of implantation: ___/___/_____
Brand and model: .....

**LEADS**

	Brand and model	Date of implantation
Atrial lead:	.....	___/___/_____
RV lead:	.....	___/___/_____
LV lead:	.....	___/___/_____

Is the system:	<input type="checkbox"/> MR conditional?			
	<input type="checkbox"/> MR non-conditional?			
Is the patient pacing dependent?	<input type="checkbox"/> YES	<input type="checkbox"/> NO	<input type="checkbox"/> NA	
Are there abandoned lead(s)?	<input type="checkbox"/> YES	<input type="checkbox"/> NO	<input type="checkbox"/> NA	
Are there epicardial lead(s)?	<input type="checkbox"/> YES	<input type="checkbox"/> NO	<input type="checkbox"/> NA	
If ICD:	<input type="checkbox"/> Primary prevention	<input type="checkbox"/> Secondary prevention or history of appropriate therapy	<input type="checkbox"/> NA	

**RECOMMENDATIONS FOR MRI**

MRI 1.5 T full body	<input type="checkbox"/>	MRI 1.5 T with thoracic exclusion	<input type="checkbox"/>
MRI 3 T full body	<input type="checkbox"/>	MRI 3 T with thoracic exclusion	<input type="checkbox"/>
MRI contraindicated	<input type="checkbox"/>		
Presence of the device specialist within premises required?	<input type="checkbox"/> YES	<input type="checkbox"/> NO	
MR-mode programming in cardiology department possible?	<input type="checkbox"/> YES	<input type="checkbox"/> NO	
Device MR programming: MR mode possible?	<input type="checkbox"/> YES	<input type="checkbox"/> NO	
<input type="checkbox"/> VOO/DOO	<input type="checkbox"/> ODO/OOO	<input type="checkbox"/> VVI/DDI	<input type="checkbox"/> ICD therapy deactivation
Reprogramming of the device after MRI necessary?	<input type="checkbox"/> YES	<input type="checkbox"/> NO	

Name and signature of cardiologist: ..... Date: \_\_\_/\_\_\_/\_\_\_\_\_

MR-conditional CEIDs

Pacing-dependant permanent bradycardia

No permanent bradycardia

ICD

VOO / DOO

ODO / OOO

VVI / DDI  
(off-label)

Deactivate ICD  
tachycardia detection

Before MRI scan

MR-mode programming before MRI scan (same day<sup>a</sup>)

Patients with paroxysmal AVB / SND : prefer VVI / DDI programming  
Patients with recent ICD therapy or unstable clinical condition : program just before MRI

During MRI scan

Device specialist available on call

Monitoring (pulse oximetry ± electrocardiogram monitoring, visual contact) by physician or qualified trained personnel  
Presence of a defibrillator and emergency material  
Physicians with the skill to perform resuscitation available immediately

After MRI scan

Device control and baseline setting restoration after MRI scan (same day<sup>a</sup>)

Patients with recent ICD therapy or unstable clinical condition : program just after MRI

Standard follow-up after MRI scan or remote monitoring  
Early follow-up if acute modification of sensing / capture / impedance

## MR-non-conditional CEIDs

Pacing-dependant permanent bradycardia

No permanent bradycardia

ICD

VOO / DOO

ODO / OOO

VVI / DDI

Deactivate ICD  
tachycardia detection

Before MRI scan

MR-mode programming before MRI scan (same day<sup>a</sup>)

Patients with paroxysmal AVB / SND : prefer VVI / DDI programming  
Patients with recent ICD therapy or unstable clinical condition : program just before MRI

During MRI scan

Device specialist available  
immediately

Device specialist on site available in case of emergency

Monitoring (pulse oximetry ± electrocardiogram monitoring, visual contact) by physician or qualified trained personnel  
Presence of a defibrillator and emergency material  
Physicians with the skill to perform resuscitation available immediately

After MRI scan

Device control and baseline setting restoration after MRI scan (same day<sup>a</sup>)

Patients with recent ICD therapy or unstable clinical condition : program just after MRI

Standard follow-up after MRI scan or remote monitoring  
Early follow-up if acute modification of sensing / capture / impedance