Using MALDI-ToF mass spectrometry to identify mushroom species – Proof of concept analysis of *Amanita* genus specimens

Raphael Piarroux¹, Frédéric Gabriel², Frédéric Grenouillet³, Patrick Collombon⁴, Philippe Louasse⁵, Martine Piarroux⁶ and Anne-Cécile Normand^{7*}

¹ LDBIO Diagnostics, Lyon, France

² Centre Hospitalier Universitaire de Bordeaux, Service de Parasitologie Mycologie, F-33000, Bordeaux, France

³ Department of Parasitology and Mycology, University Hospital of Besançon, Besançon, France

⁴ Société de Mycologie de Provence, Laboratoire de Botanique - Faculté des Sciences Saint-Charles, F-13331, Marseille, France

⁵ Société de Mycologie de France, Paris, France

⁶ Centre d'épidémiologie et de santé publique des armées (CESPA), Paris, France

⁷ AP-HP, Groupe Hospitalier La Pitié-Salpêtrière, Service de Parasitologie Mycologie, F-75013, Paris, France

* Corresponding author

Corresponding author: Anne-Cécile Normand; e-mail: annececile.normand@aphp.fr; Postal address: Hôpital Pitié Salpêtrière, 47-83 boulevard de l'Hôpital Laboratoire de Parasitologie/Mycologie, 75013 Paris, France; phone: +33 142 16 01 96; fax: +33 142 16 01 65

ABSTRACT

Food poisoning caused by toxic mushrooms, such as species in the *Amanita* genus, occurs frequently around the world. To properly treat these patients, it is important to rapidly and accurately identify the causal species. Matrix-assisted laser desorption/ionization time-of-flight (MALDI-ToF) mass spectrometry is a rapid technique that has been used in medical laboratories for the past three decades to identify bacteria, yeasts and filamentous fungi.

Matrix Assisted Laser Desorption Ionisation Time of flight mass spectrometry (MALDI-Tof MS) is a rapid method used for the past three decades to identify microorganisms. In this study, we created and internally validated a MALDI-Tof MS reference database comprising 15 *Amanita* species frequently encountered in France, and we challenged this database with 38 *Amanita* specimens from four French locations, using a free online application for MALDI-ToF spectra identifications.

Assessment of the database showed that mass spectra can be obtained by analyzing any portion of a carpophore and that all portions enabled identification of the carpophore at the species level. Most carpophores were correctly identified using our database, with the exception of specimens from the *Vaginatae* section. Decay tests also demonstrated that decayed portions (like those found in the kitchen garbage can) of *Amanita phalloides* mushrooms could be properly identified using MALDI-ToF MS.

Our findings provide important insight for toxicology laboratories that often rely on DNA sequencing to identify meal leftovers implicated in food poisoning. In future developments, this technique could also be used to detect counterfeit mushrooms by including other genera in the reference database.

LAY SUMMARY

MALDI-ToF MS is a powerful identification tool for microorganisms. We demonstrate that the technique can be applied to *Amanita* specimens. This will prevent food intoxications as a rapid and definite identification can be obtained, and it can also be used for food remnants.

INTRODUCTION

Mushrooms have been implicated in many cases of food poisoning globally. Most life-threatening poisoning cases due to ingestion of toxic mushrooms are caused by specimens belonging to the *Amanita* genus, especially *A. phalloides*, *A. virosa*, *A. verna* and *A. decipiens*. *A. phalloides* is responsible for approximately 95% of poisoning cases globally ¹. Various symptoms may develop depending on the species of mushroom ingested ^{2,3,4(p)}. Most symptoms induced by *Amanita* mushrooms are caused by amatoxins that block DNA production, thereby leading to cell death, especially in cells that require frequent renewal (e.g., liver and kidney cells). Other mycotoxins have also been implicated in cases of mushroom poisoning, such as coprine, muscarine, muscimol, ibotenic acid and gyromitrin ². The toxins in certain mushrooms can cause Phalloidic syndrome, in which the initial symptoms may include respiratory difficulties and vertigo, followed by painful vomiting, acute diarrhea and eventual severe dehydration. Subsequent symptoms may include severe toxic hepatitis, which can lead to liver destruction, followed by cerebral disorders. Patients may recover after several months of convalescence; however, up to approximately 30% of cases succumb to the intoxication ⁴⁻¹⁴. Identification of species implicated in food poisoning is important to properly treat patients, as the symptoms and appropriate treatment vary depending on the species involved ². Mushroom identification is a complex process that requires specialists, as the toxic potential of morphologically similar mushrooms is difficult to assess (e.g., *Amanita spissa* and *A. rubescens*

mushrooms are non-toxic and edible if thoroughly cooked, while phenotypically look alike *A. pantherina* carpophores can be deadly). When analyzing meal leftovers, species identification based on morphology is nearly impossible and thus requires DNA sequence analysis. However, this technique remains relatively expensive and time-consuming, requiring one to several days to obtain results, especially if food samples contain several species. Real-time PCR assays have been developed to detect the presence of some *Amanita* species in food remnants or human feces, although the method is only specific for three species (i.e., *A. phalloides*, *A. virosa* and *A. verna*)¹⁵. Consequently, toxicologists recommend treating the symptoms in food poisoning cases before attempting to identify the species, as these identification methods are often either uninformative or too time consuming¹⁶.

The MALDI-ToF mass spectrometry technique has been applied in the mycology domain to efficiently identify fungal species such as yeasts, molds and dermatophytes^{17–24}. However, to the best of our knowledge, this method has only been used once to identify mushroom species, involving a diverse panel of mushrooms from Hokkaido Island in Japan²⁵. An online identification application (MSI) has recently been created to construct reference databases of various organisms and make them available for laboratories worldwide^{26,27}. This application allows users from around the world to freely and rapidly identify spectra obtained from yeasts, filamentous fungi, dermatophytes, parasites (such as *Leishmania*), or insects (such as *Phlebotomes*) since 2017. More than 1000 references of various yeasts, filamentous fungi, sand fly and parasite species are already available in the application.

In this study, we developed a database to easily and rapidly identify mushrooms of the *Amanita* genus using MALDI-ToF mass spectrometry and the MSI online application. We present here a proof of concept that highlights the possibility to easily and rapidly identify mushrooms from a

very small amount of mushroom material, using a combination of Maldi-ToF mass spectrometry technology and the online application (MSI).

METHODS

Construction of the MSI reference database.

Carpophore collection. The local mycologist association *Société Mycologique de Provence* (SMP) in Marseille Provence (southeastern France) collected mushrooms to be included in the reference database. From September 2013 to October 2016, a total of 447 carpophores were collected in the Provence region by members of the SMP. In 2015, 3 *Amanita* carpophores were collected in the Vosges Forest (located in northeastern France). In October 2016, the local mycologist association (Cercle d'études mycologiques en Aquitaine) in Bordeaux (southwestern France) collected 34 carpophores. In 2017, 3 *Amanita* carpophores were collected in the Doubs Forest (northeastern France). In total, 487 carpophores were collected for the study.

Carpophore identification. The local mycologist associations conducted phenotype-based identification of the carpophores. Accordingly, assessment of all 484 carpophores revealed 128 genera and 366 species.

Each carpophore was also submitted for DNA sequence analysis of the ITS1-ITS2 region using the ITS1 (5'-TCCGTAGGTGAACCTGCGG-3') and ITS4 (5'-TCCTCCGCTTATTGATATGC-3') primers, according to the protocol described by Fujita et al.²⁸, with an annealing temperature of 55°C. Identification at the species level was confirmed using a phylogenetic tree and the reference identifications published by Cui et al. in 2020²⁹ (supplemental figure 1).

Construction of the mass spectrometry reference database. The *Amanita* reference database comprised spectra obtained using a Microflex LT system (Bruker Daltonics, Bremen, Germany). To analyze each *Amanita* carpophore, a 1 x 2-mm sample was extracted from 2 to 3 gills. A small sample of the stem, ring and volva portions (approximately 2 mm³) were also extracted in

115 separate tubes when available and relevant. The mushroom samples were suspended in 75%
116 ethanol HPLC to inactivate the mushroom. After a 10-minute centrifugation step at 13,000 g, the
117 hydroalcoholic solution was removed, and the pellets were suspended in at least 20 µl of 70%
118 formic acid (or enough volume to cover the pellet) (Sigma-Aldrich, France). The mushroom
119 samples were homogenized in formic acid by compressing the sample against the wall of the tube
120 and pipetting up and down. After a 5-minute incubation step that allows the cell walls to be
121 destroyed in contact with the formic acid, the same volume of acetonitrile HPLC was added
122 (VWR International S.A.S., Fontenay-sous-Bois, France), and the two reagents were mixed by
123 pipetting up and down. After 5 minutes of incubation at ambient temperature for neutralization of
124 the acid and precipitation of the proteins, the sample was centrifuged for 2 minutes at 13,000 g,
125 and 1 µl-drops of the supernatant were deposited onto the polished steel targets (MTP384, Bruker
126 Daltonics GmbH, Bremen, Germany). Each deposit was then covered with 1 µl of matrix (α -
127 cyano-4-hydroxycinnamic acid (HCCA)) (Applied BiosystemsW, Villebon sur Yvette, France).
128 Four to 10 deposits were realized for each portion of each carpophore that was sampled, and the
129 spectra were acquired using the original Bruker parameters. To compensate for the difficulties
130 encountered when acquiring stem spectra, previous studies have proposed to modify either the
131 software parameters or the extraction protocol ^{25,30}. We chose to modify the parameter in the
132 Flexcontrol software for all mushroom samples, in which the “save sum of rejected spectra”
133 parameter was selected instead of “do not save”, or “save zeroline” in the “behavior after
134 unsuccessful acquired spectrum” parameter (as previously suggested in a study on filamentous
135 fungi ³⁰). Each acceptable spectrum (i.e., visible separated peaks) was considered as a reference
136 spectrum, and a reference database was generated in the updated version of the MSI application
137 (<https://msi.happy-dev.fr/>) ²⁷.

Score determination

Each spectrum that was retained as a reference was compared against all other references using the MSI application. The comparison scores were assigned a ‘validity label’ as follows: “concordant at the species level” (reference of the same species), “concordant at the series level” (reference belonging to the same series but not the same species) or “discordant” (reference belonging to a different series). The best identification used for score determination was the first one belonging to another carpophore than the targeted spectrum.

Identification threshold

The identification threshold was determined using two graphs (figure 1 and figure 2). The first graph (figure 1) compiled, for each spectrum, three data: i/the best scores that were concordant at the species level, ii/concordant at the series level and iii/discordant. The second graph (figure 2) categorized only the best score for each spectrum according to the validity label. The threshold defined for the study was the minimum score for which there was no error of identification at the species level. After determining the threshold, any result with a score below the threshold was not considered for analysis and was labeled ‘no identification’.

Identification scores per portion

As each of the spectra used to generate the reference database was used to challenge the database, we collected, among the correct identifications at the species level, the best scores obtained against the various available portions of the same species. Thus, for a stem spectrum, we could obtain up to four scores, one against the stem portion (if not belonging to the same carpophore), one against the gills portion, one against the volva portion and one against the ring portion (from any carpophore). For all spectra, the mean scores of each portion were calculated to evaluate their identification potential.

Performance tests

1/ Using the MSI application, we compared the database against a validation sample of 9 carpophores (stem portions only) belonging to the same mycelium (i.e., collected at the same time and in the same location, belonging to the same fairy ring) as the carpophores used for the reference database. The validation sample corresponded to 6 *Amanita* species: *A. citrina*, *A. muscaria*, *A. ovoidea*, *A. pantherina*, *A. phalloides* and *A. rubescens*.

2/ Using the MSI application, we also compared the database against an external test sample of 44 carpophores (73 portions) collected during a six-year period from four regions in France (32 samples from the Vosges Forest between 2013 and 2019, 1 sample from Provence in 2016, 8 samples from the Paris area in 2019 and 3 samples from the Doubs Forest in 2017).

These carpophores were initially identified either via phenotyping or DNA sequencing. Seven *Amanita* taxa were used for these external tests (*A. citrina* (n=15), *A. rubescens* (n=11), *A. muscaria* (n=8), *A. cf. Vaginatae* section (n=7), *A. pantherina* (n=1), *A. submaculata* (n=1) and *A. junquillea* (n=1)). If the best score obtained was below the threshold defined in the first part of this study, the spectrum was considered as not identified. If the score was above the threshold, the identification result was compared to the phylogenetic identification obtained from the ITS DNA sequencing, to determine whether it was correct or not.

Decay test. A decay test was performed on 2 *A. phalloides* carpophores for several days until the sample decayed to a liquefied state at room temperature. Spectra from each portion were compared from one day to the next using the Bruker Flex Analysis software. We then generated a small-scale decay reference database with these spectra for analysis using the MSI application.

RESULTS

References. After assessing the DNA sequences and comparing the DNA results against the phenotype results, we identified a total of 41 carpophores that corresponded to the *Amanita* genus and 14 *Amanita* species. These carpophores belonged to 8 series or species subsections as

follows: the *Amanita* series (*A. muscaria*), the *Pantherina* series (*A. pantherina*, and *A. junquillea*), the *Caesareae* subsection (*A. caesarea*), the *Vaginatae* section (*A. submembranacea*), the *Validae* series (*A. rubescens* and *A. spissa*), the *Mappae* series (*A. citrina*, *A. intermedia* and *A. porphyria*), the *Ovoidea* series (*A. ovoidea* and *A. proxima*) and the *Phalloideinae* subsection (*A. phalloides* and *A. verna-var-decipiens*). Each species was represented by 1 to 6 carpophores, and most references were obtained from specimens collected in Provence (31/41). A total of 102 *Amanita* portions were used to generate 815 individual reference spectra. The details for the references of each of the 14 *Amanita* species are provided in Table 1.

Identification threshold.

We found that threshold scores above 18.64 consistently corresponded to correct identification results (with the exception of an *A. submembranacea* stem spectrum that initially matched an *A. porphyria* reference with a score of 25.05 (Figures 1 and 2).

Applying a score threshold of 20, 74% of the tested spectra were correctly identified, of which all but one were correct at the species level.

Identification scores per portion.

The identification scores were higher when comparing the same portion types (i.e., stem vs. stem) than when comparing different portion types (i.e., stem vs. gills) (Table 2). Nevertheless, all portion references enabled identification of spectra from the other portions, with the exception of the ring reference for which the identification threshold was not reached for the volva spectra and vice versa. As those two portions could be efficiently identified by the stem and gills references, we determined that stem and gill spectra are better suited to construct an efficient reference database.

Identification results.

Identification of carpophores from the same mycelium.

We analyzed 9 stems from carpophores that corresponded to the same mycelium of at least one reference in the database. Four spectra were acquired for each stem and were all correctly identified with identification scores ranging from 27.25 (*A. muscaria*) to 58.13 (*A. citrina*) (Table 3). Only 3 of 36 spectra did not reach the identification threshold and thus could not be identified.

Identification of carpophores from different mycelia.

The results obtained for the 44 carpophores (73 portions) from mycelia unrelated to other mycelia in the database are detailed in Table 4.

Among these 44 carpophores, the 7 specimens belonging to the *Vaginatae* section (5 *A. fulva*, 1 *A. umbrinolutea* and 1 unidentified) were not identified, as only *A. submembranacea* represents this section in the references database. Two of the remaining 37 carpophores were either not identified (score <20) or misidentified: 1 *A. submaculata* sample did not reach the defined threshold (this species was not represented in the database), and 1 *A. rubescens* sample from the Vosges Forest in 2019 yielded scores below the defined threshold (19.3 for the gills and 18.24 for the stem).

The other 35 carpophores were correctly identified by analyzing at least one portion. These 35 carpophores were represented by 57 portions, among which 52 (91%) were identified with a score above the defined threshold. Identification difficulties were observed with the carpophores collected in August 2019, which had very dry stems and uncharacteristic morphology due to the drought that occurred that year.

Decay test.

Two *A. phalloides* carpophores were tested over the course of three consecutive days until the samples decayed. Spectra were obtained from the stem, gill and volva portions.

On decay day 3, the spectra appeared to show fewer peaks than on day 1 or day 2 (Figure 3). Assessment of the stem profiles revealed that most peaks visible on day 1 seemed to remain over the next two days, albeit with reduced intensity. The gill and volva profiles seemed to have fewer peaks than the stem profiles between decay days 2 and 3. When assessing these spectra using the MSI application (Figure 4), we found that all references enabled identification of all decay sample spectra. However, the references obtained on day 2 enabled the most efficient identification of the spectra from decay days 1 to 3.

DISCUSSION

Identification of toxic mushrooms at the species level is important to better diagnose and treat cases of food poisoning due to toxic and potentially deadly mushroom species^{4,31}. Phenotype-based identification of mushroom species should be performed by a specialist with expertise in fungal taxonomy and morphology. Currently, DNA analysis techniques are the only widely accepted alternative to phenotype-based identification, either by sequencing various portions of specific genes or using real-time PCR assays designed to recognize a few species of interest¹⁵.

The primary objective of this study was to assess whether MALDI-ToF mass spectrometry, a rapid and less expensive technique than DNA-based methods, could be applied to identify mushrooms at the species level. We focused on the *Amanita* genus, which has been implicated in most life-threatening mushroom poisoning cases reported in France^{4,8,32}.

We successfully obtained spectra from all portions of the carpophores using MALDI-ToF mass spectrometry. To circumvent the difficulties encountered when acquiring stem spectra, we found that modification of an acquisition parameter in the Flex Control software was sufficient and that modification of the extraction protocol was unnecessary. Assessment of the database enabled us to define an identification threshold of 20, which is similar to that previously defined for pathogenic filamentous fungi²⁷. While testing the database by comparing the reference spectra

against the same spectra in a test panel, we often found that spectra obtained from different portions of the same carpophore varied but showed sufficient similarity to enable correct identification at the species level. Comparison between the ring and volva spectra revealed the lowest degree of similarity.

After defining the identification threshold, we submitted 36 spectra of carpophores belonging to the same mycelia as references for the MSI application. When considering an identification threshold of 20, identification performance was excellent with scores reaching a maximum of 76.3. When considering only the best identification results of the four spectra submitted for each mushroom sample, 100% of the samples were correctly identified.

External validation of the database was performed using carpophores collected in regions other than those represented in the reference database. The validation revealed that 80% (35/44) of the carpophores were identified by assessing at least one portion, and 71% (52/73) of the portion samples were correctly identified.

The majority of misidentifications (7 carpophores) were associated with the section *Vaginatae*, which comprises a large number of species that were not included in our reference database, and which is continuously evolving. In 2009, Neville and Poumarat³³ described over 50 species in this subsection, while new species have been described since then by other authors^{29,34,35}. In 2020, Hanss and Moreau compiled a total of 86 species in the *Vaginatae* section, using ITS DNA sequencing³⁶, and there are currently 96 provisional names recorded by Tulloss and Yang³⁷ in this section. Species-level identification of carpophores from the *Vaginatae* section using DNA databases can also be challenging, as new species have recently been phenotypically described and DNA sequences are not always available. However, when this subsection was excluded, 95% (35/37) of the carpophores and 85% (52/61) of the portions were correctly identified.

The samples collected in 2019 presented another challenge due to the drought that occurred that summer. As the stems of these mushrooms were drier than usual, many of the 2019 samples did not reach the identification threshold of 20. When only considering species represented in the database (i.e., not including *the Vaginatae section*), the identification rate was 87.5% for samples from 2013 (14/16 portions), 100% for samples from 2017 (9/9 portions) and 82% for samples from 2019 (24/29 portions). These results suggest that identification performance could be improved by including specimens collected during droughts or in particular conditions (e.g., altitude, humidity, pollution) in the database.

The decay test performed on the *A. phalloides* specimens showed that the spectra could be identified regardless of the advanced state of decay. As the references obtained on a specific day enabled identification of spectra obtained from the day before and the day after, it is unnecessary to generate a database of decayed sample references obtained each day. As the timeframe of symptom development varies depending on the species implicated, mushroom databases should consider at least one reference of each portion every two days during the decay phase. It is important that the protein profile of the mushroom portion most likely to be recovered from the garbage does not markedly evolve in the decay process. MALDI-ToF mass spectrometry should help toxicologist to identify the cause of the poisoning regardless of the stage of decay¹⁶.

In conclusion, this study shows that mushrooms, and in this case the *Amanita* genus, can be rapidly and consistently identified at the species level using MALDI-ToF mass spectrometry and the MSI application, even on decayed samples. This study provides important insight for toxicology laboratories that often utilize DNA sequencing to identify pathogenic mushrooms and diagnose patients. In future developments, such MALDI-ToF-based identification approaches could also be used to detect counterfeit mushrooms.

Acknowledgements

We would like to thank the following collaborators: Magali LABADIE, Arnaud COURTOIS and Françoise PENOUIL from the Bordeaux University Hospital Anti-Poison Center; Michel PUJOL, Jacques BOYER, Jacques BECK-CECCALDI and other members of CEMA for their expertise and support during the friendly sampling campaign in Cap-Ferret; Patrick Collombon, Francis Fouchier, Serge Poumarat and members of the Société de Mycologie de Provence for providing specimens from Provence and morphological identification results; the Société de Mycologie de France for providing specimens from the Paris area; and Martine Palous from the Mycology Laboratory at the Pitié Salpêtrière Hospital who collected carpophores while running in the Fontainebleau Forest. Finally, the authors thank Sandy Moore for proofreading the English version of this manuscript.

Conflicts of Interest : None

REFERENCES

1. Neville P, Poumarat S. *Amaniteae. Amanita, Limacella & Torrendia*. Edizioni Candusso.; 2004.
2. Cassier A. Les intoxications par les champignons. Published 2014. <http://www.champinews.fr/pages/intox.pdf>
3. Garcia J, Costa VM, Carvalho A, et al. Amanita phalloides poisoning: Mechanisms of toxicity and treatment. *Food Chem Toxicol*. 2015;86:41-55. doi:10.1016/j.fct.2015.09.008
4. Sinno-Tellier S, Bruneau C, Daoudi J, Greillet C, Verrier A, Bloch J. Surveillance nationale des intoxications alimentaires par des champignons : bilan des cas rapportés au réseau des centres antipoison de 2010 à 2017 en France métropolitaine / National surveillance of food poisoning by mushrooms: cases reported to the network of Poison Control Centres from 2010 to 2017. http://beh.santepubliquefrance.fr/beh/2019/33/2019_33_1.html. Published online December 2019:13.
5. Brandão JL, Pinheiro J, Pinho D, et al. Mushroom poisoning in Portugal. *Acta Médica Portuguesa*. 2011;24(0):269-278. doi:10.20344/amp.1491

- 332 6. Fantozzi R, Ledda F, Caramelli L, et al. Clinical findings and follow-up evaluation of an
333 outbreak of mushroom poisoning--survey of *Amanita phalloides* poisoning. *Klin*
334 *Wochenschr.* 1986;64(1):38-43. doi:10.1007/BF01721579
- 335 7. Giannini L, Vannacci A, Missanelli A, et al. Amatoxin poisoning: A 15-year retrospective
336 analysis and follow-up evaluation of 105 patients. *Clinical Toxicology.* 2007;45(5):539-542.
337 doi:10.1080/15563650701365834
- 338 8. Jaeger A, Jehl F, Flesch F, Sauder P, Kopferschmitt J. Kinetics of amatoxins in human
339 poisoning: Therapeutic implications. *Journal of Toxicology: Clinical Toxicology.*
340 1993;31(1):63-80. doi:10.3109/15563659309000374
- 341 9. Jander S, Bischoff J, Woodcock BG. Plasmapheresis in the treatment of *Amanita phalloides*
342 poisoning: II. A review and recommendations. *Ther Apher.* 2000;4(4):308-312.
343 doi:10.1046/j.1526-0968.2000.004004308.x
- 344 10. Schenk-Jaeger, Katharina M., Rauber-Lüthy C, Bodmer M, Kupferschmidt H, Kullak-Ublick
345 GA, Ceschi A. Mushroom poisoning: A study on circumstances of exposure and patterns of
346 toxicity | Elsevier Enhanced Reader. doi:10.1016/j.ejim.2012.03.014
- 347 11. Moroni F, Fantozzi R, Masini E, Mannaioni PF. A trend in the therapy of *Amanita phalloides*
348 poisoning. *Arch Toxicol.* 1976;36(2):111-115. doi:10.1007/BF00351969
- 349 12. Mowry JB, Spyker DA, Cantilena LR, Bailey JE, Ford M. 2012 Annual Report of the American
350 Association of Poison Control Centers' National Poison Data System (NPDS): 30th Annual
351 Report. *Clinical Toxicology.* 2013;51(10):949-1229. doi:10.3109/15563650.2013.863906
- 352 13. Olson KR, Pond SM, Seward J, Healey K, Woo OF, Becker CE. *Amanita phalloides*-type
353 mushroom poisoning. *West J Med.* 1982;137(4):282-289.
- 354 14. Pinson CW, Daya MR, Benner KG, et al. Liver transplantation for severe *Amanita phalloides*
355 mushroom poisoning. *Am J Surg.* 1990;159(5):493-499. doi:10.1016/s0002-9610(05)81254-
356 1
- 357 15. Gausterer C, Penker M, Krisai-Greilhuber I, Stein C, Stimpfl T. Rapid genetic detection of
358 ingested *Amanita phalloides*. *Forensic Sci Int Genet.* 2014;9:66-71.
359 doi:10.1016/j.fsigen.2013.11.002
- 360 16. Eren SH, Demirel Y, Ugurlu S, Korkmaz I, Aktas C, Güven FMK. Mushroom poisoning:
361 retrospective analysis of 294 cases. *Clinics (Sao Paulo).* 2010;65(5):491-496.
362 doi:10.1590/S1807-59322010000500006
- 363 17. Normand A-C, Cassagne C, Ranque S, et al. Assessment of various parameters to improve
364 MALDI-TOF MS reference spectra libraries constructed for the routine identification of
365 filamentous fungi. *BMC Microbiol.* 2013;13:76. doi:10.1186/1471-2180-13-76

- 366 18. Normand A-C, Gabriel F, Riat A, et al. Optimization of MALDI-ToF mass spectrometry for
367 yeast identification: a multicenter study. *Med Mycol*. Published online October 3, 2019.
368 doi:10.1093/mmy/myz098
- 369 19. L'Ollivier C, Cassagne C, Normand A-C, et al. A MALDI-TOF MS procedure for clinical
370 dermatophyte species identification in the routine laboratory. *Med Mycol*. 2013;51(7):713-
371 720. doi:10.3109/13693786.2013.781691
- 372 20. da Cunha KC, Riat A, Normand A-C, et al. Fast identification of dermatophytes by MALDI-
373 TOF/MS using direct transfer of fungal cells on ground steel target plates. *Mycoses*.
374 Published online May 15, 2018. doi:10.1111/myc.12793
- 375 21. De Respinis S, Monnin V, Girard V, et al. Matrix-assisted laser desorption ionization-time of
376 flight (MALDI-TOF) mass spectrometry using the Vitek MS system for rapid and accurate
377 identification of dermatophytes on solid cultures. *J Clin Microbiol*. 2014;52(12):4286-4292.
378 doi:10.1128/JCM.02199-14
- 379 22. Lau AF, Drake SK, Calhoun LB, Henderson CM, Zelazny AM. Development of a clinically
380 comprehensive database and a simple procedure for identification of molds from solid
381 media by matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J*
382 *Clin Microbiol*. 2013;51(3):828-834. doi:10.1128/JCM.02852-12
- 383 23. Bader O, Weig M, Taverne-Ghadwal L, Lugert R, Gross U, Kuhns M. Improved clinical
384 laboratory identification of human pathogenic yeasts by matrix-assisted laser desorption
385 ionization time-of-flight mass spectrometry. *Clin Microbiol Infect*. 2011;17(9):1359-1365.
386 doi:10.1111/j.1469-0691.2010.03398.x
- 387 24. Buchan BW, Ledebor NA. Advances in identification of clinical yeast isolates by use of
388 matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin*
389 *Microbiol*. 2013;51(5):1359-1366. doi:10.1128/JCM.03105-12
- 390 25. Sugawara R, Yamada S, Tu Z, et al. Rapid and reliable species identification of wild
391 mushrooms by matrix assisted laser desorption/ionization time of flight mass spectrometry
392 (MALDI-TOF MS). *Analytica Chimica Acta*. 2016;934:163-169.
393 doi:10.1016/j.aca.2016.05.056
- 394 26. Lachaud L, Fernández-Arévalo A, Normand A-C, et al. Identification of *Leishmania* by
395 Matrix-Assisted Laser Desorption Ionization-Time of Flight (MALDI-TOF) Mass
396 Spectrometry Using a Free Web-Based Application and a Dedicated Mass-Spectral Library. *J*
397 *Clin Microbiol*. 2017;55(10):2924-2933. doi:10.1128/JCM.00845-17
- 398 27. Normand AC, Becker P, Gabriel F, et al. Validation of a New Web Application for
399 Identification of Fungi by Use of Matrix-Assisted Laser Desorption Ionization-Time of Flight
400 Mass Spectrometry. *J Clin Microbiol*. 2017;55(9):2661-2670. doi:10.1128/JCM.00263-17

28. Fujita S-I, Senda Y, Nakaguchi S, Hashimoto T. Multiplex PCR Using Internal Transcribed Spacer 1 and 2 Regions for Rapid Detection and Identification of Yeast Strains. *Journal of Clinical Microbiology*. 2001;39(10):3617-3622. doi:10.1128/JCM.39.10.3617-3622.2001
29. Cui Y-Y, Cai Q, Tang L-P, Liu J-W, Yang ZL. The family Amanitaceae: molecular phylogeny, higher-rank taxonomy and the species in China. *Fungal Diversity*. 2018;91(1):5-230. doi:10.1007/s13225-018-0405-9
30. Normand A-C, Cassagne C, Gautier M, et al. Decision criteria for MALDI-TOF MS-based identification of filamentous fungi using commercial and in-house reference databases. *BMC Microbiol*. 2017;17(1):25. doi:10.1186/s12866-017-0937-2
31. Unknown. Mushroom poisoning. *The Lancet*. 1980;316(8190):351-352. doi:10.1016/S0140-6736(80)90346-3
32. French LK, Hendrickson RG, Horowitz BZ. Amanita phalloides poisoning. *Clin Toxicol (Phila)*. 2011;49(2):128-129. doi:10.3109/15563650.2011.557663
33. Neville P, Poumarat S. *Fungi non Delineati 51-52: Quelques Espèces Nouvelles ou Mal Delimitées d'Amanita de la Sous-section Vaginatae [Some New or Poorly Delimited Species of Amanita of Subsection Vaginatae]*. Edizioni Candusso.; 2009.
34. Loizides M, Bellanger J-M, Yiangou Y, Moreau P-A. Preliminary phylogenetic investigations into the genus Amanita (Agaricales) in Cyprus, with a review of previous records and poisoning incidents. 2018;XXXVII:201-218.
35. Ševčíková H, Hanss J-M, Moreau P-A. Amanita vladimirii (Amanitaceae, Agaricales), a new European species in section Vaginatae. *Phytotaxa*. 2021;482(2):159-172. doi:10.11646/phytotaxa.482.2.4
36. Hanss JM, Moreau PA. Une Révision des Amanites «vaginées»(Amanita sect. Vaginatae) en Europe, 1re partie: quelques Amanites argentées. *Bulletin de la Société mycologique de France*. 2020;133(1-2):67-141.
37. Tulloss RE, Yang Z-L. section Vaginatae - Amanitaceae.org - Taxonomy and Morphology of Amanita and Limacella. Accessed February 17, 2021. http://www.amanitaceae.org/?page_id=2269&sort=IGFuZCBsaS5sYWJlbD0nbm9tLiBwcm92Licg

FIGURES LEGENDS

Figure 1: Best concordant at the species level, concordant at the series level and discordant scores for each spectrum.

Figure 2: Repartition of the best score for each spectrum.

Figure 3: Evolution of the decay spectra for each portion for the two *A. phalloides* carpophores.

Figure 4: Score distribution for the various decay days (white boxes: spectra realized at decay day 1; gray boxes: spectra realized at decay day 2; black boxes: spectra realized at decay day 3). The lower and upper portions of the box represent the lower and upper quartiles, respectively. The line represents the median value. The ends of the whiskers represent the lowest datum included in the 1.5 interquartile range of the lower quartile and the highest datum included in the 1.5 interquartile range of the upper quartile. Outlier values are represented by circles. For each spectrum, scores noted correspond to the identification against the same specimen but at different decay days. Therefore, for each spectrum, three scores were used: decay day 1, decay day 2 and decay day 3.

Supplemental Figure 1: Phylogeny using ITS sequences of the specimens collected in this study, using the maximum likelihood (ML) analysis, and 1000 bootstraps. Bootstrap values are represented by stars, and values below 70 are not shown. Sequences from Cui et al.²⁹ are highlighted in red and displayed in italic. Specimens used as MS references are displayed in bold.

455 **Table 1:** Details of the references obtained for each of the 15 *Amanita* species

Species	Carpophore origin	NCBI accession number for ITS sequences	Number of portions / number of stem references	Number of portions / number of gill references	Number of portions / number of volva references	Number of portions / number of ring references	Number of portions / number of total references	Harvest year	Decay test
<i>A. caesarea</i>	Provence (1)	MW589070	1 / 10	1 / 10	1 / 10	1 / 10	4 / 40	2013 (1)	No
<i>A. citrina</i>	Provence (6) Vosges (1)	MW589071* MW589072 MW589073 MW589074 MW589075	7 / 52	7 / 52	0 / 0	2 / 20	16 / 124	2013 (2), 2014 (1), 2015 (1), 2016 (3)	No
<i>A. intermedia</i>	Provence (1)	MW589076	1 / 10	1 / 10	0 / 0	0 / 0	2 / 20	2013 (1)	No
<i>A. junquillea</i>	Provence (3)	MW589077 MW589078*	3 / 18	3 / 18	0 / 0	0 / 0	6 / 36	2014 (1), 2016 (2)	No
<i>A. muscaria</i>	Provence (3) Bordeaux (1) Vosges (1)	MW589079 MW589080 MW589081 No sequence available MW589082	4 / 34	5 / 44	1 / 4	1 / 10	11 / 92	2013 (2), 2014 (1), 2015 (1), 2016 (1)	No
<i>A. ovoidea</i>	Provence (2)	No sequence available	2 / 19	2 / 20	0 / 0	0 / 0	4 / 39	2013 (2)	No
<i>A. pantherina</i>	Provence (1) Bordeaux (2)	MW589083 MW589126*	3 / 18	3 / 18	0 / 0	0 / 0	6 / 36	2013 (1), 2016 (2)	No
<i>A. phalloides</i>	Provence (4) Bordeaux (2)	MW589084 MW589085 MW589086 MW589087 MW589127*	10 / 88	10 / 88	6 / 60	0 / 0	26 / 236	2013 (3), 2014 (1), 2016 (2)	Day 1, day 2 and day 3
<i>A. porphyria</i>	Provence (3)	MW589088*	3 / 12	3 / 12	0 / 0	0 / 0	6 / 24	2016 (3)	No
<i>A. proxima</i>	Provence (1)	MW589089	1 / 10	1 / 10	0 / 0	0 / 0	2 / 20	2013 (1)	No
<i>A. rubescens</i>	Provence (2) Bordeaux (2)	MW589091 MW589092 MW589128*	4 / 28	4 / 28	1 / 4	0 / 0	9 / 60	2014 (2), 2016 (2)	No
<i>A. spissa</i>	Provence (1)	MW589090	1 / 10	1 / 10	0 / 0	0 / 0	2 / 20	2013 (1)	No
<i>A. submembranacea</i>	Provence (1) Vosges (1)	MW589094 MW589093	2 / 14	2 / 14	0 / 0	0 / 0	4 / 28	2015 (1), 2016 (1)	No
<i>A. verna</i> var <i>decipiens</i>	Provence (2)	MW589095	2 / 20	2 / 20	0 / 0	0 / 0	4 / 40	2016 (2)	No

456

457 *several different carpophores from the same mycelium incorporated as references, only one

458 sequenced.

Table 2: Mean and standard deviation of the scores obtained for each spectrum against a reference of each portion of the same species. The number of spectra that yielded a correct identification against the specified portion is indicated in parentheses.

Portion spectra	Reference portion	Ring mean score +st dev	Stem mean score +st dev	Gill mean score +st dev	Volva mean score +st dev
Ring (n=40)		78.13 ± 11.61 (n=40)	29.77 ± 6.45 (n=40)	29.23 ± 6.22 (n=39)	16.54 ± 4.11 (n=20)
Stem (n=343)		21.96 ± 8.73 (n=94)	76.15 ± 13.39 (n=323)	29.95 ± 9.85 (n=302)	27.83 ± 12.10 (n=136)
Gill (n=354)		23.63 ± 7.69 (n=103)	29.07 ± 9.60 (n=301)	70.11 ± 16.04 (n=334)	22.51 ± 9.81 (n=150)
Volva (n=78)		13.51 ± 5.78 (n=14)	34.31 ± 9.99 (n=74)	27.33 ± 8.61 (n=73)	67.69 ± 15.36 (n=78)

466 **Table 3:** Results for nine carpophores belonging to the same mycelium of at least one other
467 reference in the database.

Carpophore sample	Portion	Identified species	Maximum score	Validation
Provence2013-M124_ <i>A. muscaria</i>	stem	<i>Amanita muscaria</i>	55.26	Correct
Provence2013-M127_ <i>A. citrina</i>	stem	<i>Amanita citrina</i>	76.3	Correct
Provence2013-M132_ <i>A. citrina</i>	stem	<i>Amanita citrina</i>	65.53	Correct
Provence2013-M004_ <i>A. intermedia</i>	stem	<i>Amanita intermedia</i>	42.63	Correct
Provence2013-M015_ <i>A. muscaria</i>	stem	<i>Amanita muscaria</i>	29.28	Correct
Provence2013-M166_ <i>A. pantherina</i>	stem	<i>Amanita pantherina</i>	54.42	Correct
Provence2013-M171_ <i>A. rubescens</i>	stem	<i>Amanita rubescens</i>	53.3	Correct
Provence2013-M174_ <i>A. ovoidea</i>	stem	<i>Amanita ovoidea</i>	29.3	Correct
Provence2013-M180_ <i>A. phalloides</i>	stem	<i>Amanita phalloides</i>	47.15	Correct

468

469

470 **Table 4:** Results for the 44 carpophores from mycelia unrelated to other mycelia in the database.

471 Identification results with scores below 20 were not considered due to the high likelihood of

472 misidentification.

Sample-DNA identification	NCBI accession number for ITS sequences	Portion	MSI identification	Maximum score	Validation
Doubs2017-A_A. <i>citrina</i>	No sequence available	stem	<i>Amanita citrina</i>	33.9	Correct
Doubs2017-B_A. <i>muscaria</i>	No sequence available	stem	<i>Amanita muscaria</i>	28.6	Correct
Doubs2017-C_A. <i>rubescens</i>	No sequence available	stem	<i>Amanita rubescens</i>	25.3	Correct
Paris2019-A_A. <i>citrina</i>	MW589096	ring	<i>Amanita citrina</i>	21.1	Correct
		stem	<i>Amanita citrina</i>	23.2	Correct
		gill	<i>Amanita citrina</i>	30.4	Correct
		volva	Unidentified	0	No identification
Paris2019-B_A. <i>af. Vaginatae section</i>	No ITS sequence available	stem	Unidentified	0	No identification
		gill	Unidentified	0	No identification
Paris2019-C_A. <i>muscaria</i>	MW589097	stem	<i>Amanita muscaria</i>	48.2	Correct
		gill	<i>Amanita muscaria</i>	52.1	Correct
Paris2019-D_A. <i>pantherina</i>	No ITS sequence available	stem	Unidentified	0	No identification
		gill	<i>Amanita pantherina</i>	33.3	Correct
Paris2019-E_A. <i>A. rubescens</i>	No ITS sequence available	ring	<i>Amanita rubescens</i>	31.4	Correct
		stem	<i>Amanita rubescens</i>	30.1	Correct
		gill	<i>Amanita rubescens</i>	37.2	Correct
Paris2019-F_A. <i>submaculata</i>	MW589098	stem	Unidentified	0	No identification
		gill	Unidentified	0	No identification
Paris2019-G_A. <i>rubescens</i>	MW589099	stem	<i>Amanita rubescens</i>	32.3	Correct
		gill	<i>Amanita rubescens</i>	32.4	Correct
Paris2019-H_A. <i>citrina</i>	No sequence available	stem	<i>Amanita citrina</i>	36.8	Correct
		gill	<i>Amanita citrina</i>	49.1	Correct
Provence2016-M338_A. <i>muscaria</i>	MW589100	stem	<i>Amanita muscaria</i>	46.7	Correct
		gill	<i>Amanita muscaria</i>	46.4	Correct
Vosges2013-C_A. <i>rubescens</i>	MW589101	stem	Unidentified	26.9	Correct
		gill	<i>Amanita rubescens</i>	31.5	Correct
Vosges2013-D_A. <i>fulva</i>	MW589102	stem	Unidentified	0	No identification
		gill	Unidentified	0	No identification
Vosges2013-E_A. <i>muscaria</i>	MW589103	stem	<i>Amanita muscaria</i>	52.9	Correct
		gill	<i>Amanita muscaria</i>	43.7	Correct
Vosges2013-F_A. <i>citrina</i>	MW589104*	stem	<i>Amanita citrina</i>	40.7	Correct
Vosges2013-F1_A. <i>citrina</i>		gill	<i>Amanita citrina</i>	46.5	Correct
Vosges2013-F2_A. <i>citrina</i>		gill	<i>Amanita citrina</i>	58.4	Correct
		gill	<i>Amanita citrina</i>	43.6	Correct
Vosges2013-G_A. <i>citrina</i>	MW589105*	stem	<i>Amanita citrina</i>	23.1	Correct
		gill	<i>Amanita citrina</i>	457	Correct
Vosges2013-G1_A. <i>citrina</i>		gill	<i>Amanita citrina</i>	30.5	Correct
Vosges2013-G2_A. <i>citrina</i>		gill	Unidentified	29.3	Correct
Vosges2013-H1_A. <i>junquillea</i>	MW589106	stem	Unidentified	0	No identification
		gill	<i>Amanita junquillea</i>	21.6	Correct
Vosges2013-H2_A. <i>citrina</i>	MW589107	stem	Unidentified	36.1	Correct
		gill	<i>Amanita citrina</i>	19	No identification
Vosges2017-A_A. <i>citrina</i>	MW589108	gill	<i>Amanita citrina</i>	49.3	Correct
Vosges2017-B_A. <i>citrina</i>	MW589109	gill	<i>Amanita citrina</i>	46.6	Correct
Vosges2017-C_A. <i>citrina</i>	MW589110	gill	<i>Amanita citrina</i>	42	Correct
Vosges2017-D_A. <i>muscaria</i>	MW589111	gill	<i>Amanita muscaria</i>	46.7	Correct
Vosges2017-E_A. <i>muscaria</i>	MW589112	gill	<i>Amanita muscaria</i>	55	Correct
Vosges2017-F_A. <i>muscaria</i>	MW589113	gill	<i>Amanita muscaria</i>	48.1	Correct
Vosges2017-G_A. <i>fulva</i>	MW589114	gill	Unidentified	0	No identification

Vosges2017-H_A. <i>umbrinolutea</i>	MW589115	gill	Unidentified	0	No identification
Vosges2017-I_A. <i>rubescens</i>	No sequence available	gill	<i>Amanita rubescens</i>	36.9	Correct
Vosges2017-J_A. <i>rubescens</i>	No sequence available	gill	<i>Amanita rubescens</i>	39.8	Correct
Vosges2017-K_A. <i>rubescens</i>	No sequence available	gill	<i>Amanita rubescens</i>	28.5	Correct
Vosges2019-A_A. <i>fulva</i>	MW589116	stem	Unidentified	0	No identification
		gill	Unidentified	0	No identification
Vosges2019-B_A. <i>rubescens</i>	MW589117	stem	<i>Amanita rubescens</i>	18.3	No identification
		gill	<i>Amanita rubescens</i>	41.3	Correct
Vosges2019-C_A. <i>fulva</i>	MW589118	stem	Unidentified	0	No identification
		gill	Unidentified	0	No identification
Vosges2019-D_A. <i>rubescens</i>	MW589119	stem	Unidentified	17.1	No identification
		gill	<i>Amanita rubescens</i>	19.3	No identification
Vosges2019-E_A. <i>citrina</i>	MW589120	stem	<i>Amanita citrina</i>	31.7	Correct
		gill	<i>Amanita citrina</i>	39.2	Correct
Vosges2019-F_A. <i>citrina</i>	MW589121	stem	<i>Amanita citrina</i>	45	Correct
		gill	<i>Amanita citrina</i>	49.3	Correct
Vosges2019-G_A. <i>fulva</i>	MW589122	stem	Unidentified	0	No identification
		gill	Unidentified	0	No identification
Vosges2019-H_A. <i>muscaria</i>	MW589123	stem	<i>Amanita muscaria</i>	47.9	Correct
		gill	<i>Amanita muscaria</i>	44	Correct
Vosges2019-I_A. <i>rubescens</i>	MW589124	stem	<i>Amanita rubescens</i>	24.8	Correct
		gill	<i>Amanita rubescens</i>	40.3	Correct
Vosges2019-J_A. <i>rubescens</i>	MW589125	stem	<i>Amanita rubescens</i>	35.7	Correct
		gill	<i>Amanita rubescens</i>	34.7	Correct

473

474 *several different carpophores from the same mycelium incorporated as references, only one

475 sequenced.

Amanita phalloides M150 Stem portion

Amanita phalloides M150 Gills portion

Amanita phalloides M150 Volva portion

Amanita phalloides M160 Stem portion

Amanita phalloides M160 Gills portion

Amanita phalloides M160 Volva portion

Species
<i>A.caesarea</i>
<i>A.citrina</i>
<i>A.intermedia</i>
<i>A.porphyrta</i>
<i>A.muscaria</i>
<i>A.pantherina</i>
<i>A.junquillea</i>
<i>A.rubescens</i>
<i>A.spissa</i>
<i>A.submaculata</i>
<i>A.phalloides</i>
<i>A.verna</i>
<i>A.virosa</i>
<i>A.fulva</i>
<i>A.submembranacea</i>
<i>A.umbrinolutea</i>
<i>A.proxima</i>